

R&S®NRP-Sanitizer

Release Notes

© 2019 Rohde & Schwarz GmbH & Co. KG
Muehldorfstr. 15, 81671 Munich, Germany
Phone: +49 89 41 29 - 0
Fax: +49 89 41 29 12 - 164
E-mail: <mailto:info@rohde-schwarz.com>
Internet: <http://www.rohde-schwarz.com>

Subject to change

R&S® is a registered trademark of Rohde & Schwarz GmbH & Co. KG. Trade names are trademarks of the owners.

Version 2.5.3 | R&S®NRP-Sanitizer

The software makes use of several valuable open source software packages. For information, see the "Open Source Acknowledgment" provided with the product.

The following abbreviations are used throughout this document: R&S®NRP-Sanitizer is abbreviated as R&S NRP-Sanitizer.

PAD-TM: 3574-3288-02/03.00/C/1/EN

Contents

1	Information on the Current Version and History.....	3
1.1	New Functions	3
1.2	Modified Functions	3
1.3	Improvements	3
1.4	Known Issues.....	4
2	Modifications to the Documentation	5
3	Customer Support	6

1 Information on the Current Version and History

1.1 New Functions

The following table lists the new functions and indicates the version in which the new function was introduced:

New Function of Firmware V2.5.3:

Version	Function
2.5.3	No new functions added

1.2 Modified Functions

The following table lists the modified functions and indicates the version in which the modification was carried out:

Modifications of Firmware V2.5.3:

Version	Function
2.5.3	Fixed Calibration Data Set Handling for NRQ

Modifications of Firmware V2.5.2:

Version	Function
2.5.2	Added support for NRT2, NRX and NRQ

Modifications of Firmware V2.2.5:

Version	Function
2.2.5	Re-implementation of sanitization for NRPxxS[N], NRPxxA[N], NRPxxT[N] power sensors (based on old sanitizer app v2.1.7)

1.3 Improvements

The following tables list the improvements and indicate since which version the issue could be observed:

Improvements of Firmware V2.5.3:

since	Function
2.5.3	Fixed exception in sanitization of NRQ

1.4 Known Issues

The following tables list the known issues and indicate since which version the issue could be observed:

Known issues of Firmware V2.5.3:

since	Function
2.5.3	None

2 Modifications to the Documentation

The current documentation is up-to-date.

3 Customer Support

Technical support – where and when you need it

For quick, expert help with any Rohde & Schwarz equipment, contact one of our Customer Support Centers. A team of highly qualified engineers provides telephone support and will work with you to find a solution to your query on any aspect of the operation, programming or applications of Rohde & Schwarz equipment.

Up-to-date information and upgrades

To keep your instrument up-to-date and to be informed about new application notes related to your instrument, please send an e-mail to the Customer Support Center stating your instrument and your wish. We will take care that you will get the right information.

Europe, Africa, Middle East

Phone +49 89 4129 12345

customersupport@rohde-schwarz.com

North America

Phone 1-888-TEST-RSA (1-888-837-8772)

customer.support@rsa.rohde-schwarz.com

Latin America

Phone +1-410-910-7988

customersupport.la@rohde-schwarz.com

Asia/Pacific

Phone +65 65 13 04 88

customersupport.asia@rohde-schwarz.com

China

Phone +86-800-810-8828 / +86-400-650-5896

customersupport.china@rohde-schwarz.com