



Certificate of Compliance

Certificate: 2571181

Master Contract: 190420

Project: 80062486

Date Issued: 2020-12-21

Issued To: Rohde & Schwarz Gmbh & Co. KG Abt. 6FRK
Postfach 80 14 69
Muehldorfstrasse 15
Muenchen, Bayern, 81614
Germany

Attention: Max Weissenbacher

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: *Andreas Kohler*
Andreas Kohler



PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Certified to US Standards

Vector Signal Generator

SMW Series, Models 1412.0000X02 [SMW200A], 1440.8002X02 [SMM100A], cord connected, rated AC 100...240 V / 50...60 Hz / 400 Hz, 7.3...4.6 A or AC 100...240 V / 50...60 Hz, AC 100...120 V / 400 Hz, 8.9...4.9 A; Pollution Degree 2, Overvoltage Category II.

Broadcast Test Center

BTC Series, Model 2114.3000X02 [BTC]; cord connected, rated AC 100...240 V / 50...60 Hz / 400 Hz, 7.3...4.6 A or AC 100...240 V / 50...60 Hz, AC 100...120 V / 400 Hz, 8.9...4.9 A; Pollution Degree 2, Overvoltage Category II

Note: Type designation structure: nnnn.nnnn X nn (where n = 0 to 9 and X may be "." or "K").



Certificate: 2571181
Project: 80062486

Master Contract: 190420
Date Issued: 2020-12-21

APPLICABLE REQUIREMENTS

- | | |
|--|--|
| CAN/CSA-C22.2 No. 61010-1-12,
UPD1:2015, UPD2:2016, AMD1:2018 | - Safety Requirements for Electrical Equipment for Measurement, Controls, and Laboratory Use, Part 1: General Requirements |
| UL 61010-1 Third Edition (2012),
AMD1:2018 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |

CONDITIONS OF ACCEPTABILITY

- (1) The mains supply cord set or power supply cord with plug (< 3 m) provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.