



# Certificate of Compliance

Certificate: 70002992

Master Contract: 190420

Project: 80059047

Date Issued: 2020-11-26

Issued To: Rohde & Schwarz Gmbh & Co. KG Abt. 6FRK  
Postfach 80 14 69  
Muehldorfstrasse 15  
Muenchen, Bayern, 81614  
Germany

Attention: Max Weissenbacher

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*

Issued by: *Christoph Choudhuri*  
Christoph Choudhuri



## **PRODUCTS**

CLASS – 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS – 3631 86 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Certified to US Standards

**Radio test set CMA Series**, Model 1173.2000X18, type designation structure: nnnn.nnnn X nn (where n = 0 to 9 and X may be "." or "K"), all models rated AC 100 - 240 V, 50 - 60 / 400 Hz, 10 - 4.2 A (for use with internal power supply) or DC 10 - 30 V, max. 27 A, if DC option is used, Pollution Degree 2, Overvoltage Category II, temperature range 5...50 °C; Operation altitude of 4600 m.



**Certificate:** 70002992  
**Project:** 80059047

**Master Contract:** 190420  
**Date Issued:** 2020-11-26

---

**APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-12,  
UPD1:2015, UPD2:2016, AMD1:2018

- Safety Requirements for Electrical Equipment for Measurement, Controls, and Laboratory Use, Part 1: General Requirements

UL 61010-1 Third Edition (2012),  
AMD1:2018

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

**CONDITIONS OF ACCEPTABILITY**

- (1) The mains supply cord set or power supply cord with plug (< 3 m) provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold