



Certificate of Compliance

Certificate: 0002376094 (114196_0_000)

Master Contract: 190420

Project: 70067808

Date Issued: 2016-05-12

Issued to: Rohde & Schwarz Gmbh & Co. KG
Postfach 80 14 69
Muehldorfstrasse 15
Muenchen/Bayern, 81614
GERMANY
Attention: Max Weissenbacher

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: Bertram Mihm
Bertram Mihm

PRODUCTS

CLASS - C363106 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS - C363186 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE-Certified to US Standards

RF Source SGMA Series, Models 1416.0505X02 [SGS100A], 1418.2005X02 [SGU100A], 1419.4501X02 [SGT100A], Type designation structure: nnnn.nnnn X nn (where n = 0 to 9 and X may be "." or "K"), cord connected, rated AC 100 - 240 V, 50 - 60 Hz, max. 1.7 A, Pollution Degree 2, Installation Category II.

APPLICABLE REQUIREMENTS

- | | | |
|-------------------------------------|---|--|
| CAN/CSA-C22.2 No. 61010-1-12 | - | Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements |
| UL Std. No. 61010-1 (Third Edition) | - | Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |

CONDITIONS OF ACCEPTABILITY

- (1) The mains supply cord set or power supply cord with plug (< 3 m) provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.