ROHDE&SCHWARZ





This is to certify that:

Equipment Type	Stock No.	Designation
FPL1003	1304.0004n03	SPECTRUM ANALYZER 3GHZ
FPL1007	1304.0004n07	7GHZ SPEKTRUM ANALYZER
FPL1014	1304.0004n14	SPEKTRUM ANALYZER 14GHZ
FPL1026	1304.0004n26	SPEKTRUM ANALYZER 26GHZ
	n = "." or "K"	

complies with the provisions of the Directive of the Council of the European Union on the approximation of the laws of the Member States.

EMC Directive (2014/30/EU)

Harmonized standards applied: EN 61326-1:2013

EN 55011:2016 + A1:2017 + A11:2020

Low Voltage Directive (2014/35/EU)

Harmonized standards applied: EN 61010-1:2010+A1:2019+A1:2019/AC:2019

Restriction of the Use of Certain Hazardous Substances Directive (2011/65/EU)

Harmonized standards applied: EN IEC 63000:2018

The immunity test requirements for equipment intended to be used in industrial electromagnetic environment (acc. to EN 61326-1, Table 2) have been used as a basis.

The limits of radio interference for Group I, Class A equipment (acc. to EN 55011) have been used as a basis.

This declaration of conformity is issued under the sole responsibility of the manufacturer.

ROHDE & SCHWARZ GmbH & Co. KG Mühldorfstr. 15, D-81671 Munich

Corporate Quality Management GF-QP / Robert Lambertus EN MH Page 1 / 1