



Certificate of Compliance

Certificate: 70002782

Master Contract: 190420

Project: 80038448

Date Issued: 2020-04-14

Issued To: Rohde & Schwarz Gmbh & Co. KG Abt. 6FRK
Postfach 80 14 69
Muehldorfstrasse 15
Muenchen, Bayern, 81614
Germany

Attention: Max Weissenbacher

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: *Andreas Kohler*
Andreas Kohler



PRODUCTS

CLASS 363106 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 363186 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Certified to US Standards

Ultra-compact drive test scanner TSME Series, Model 1514.6520X02 [TSME], 4900.0004X02 [TSME6], 4900.0004X52 [TSME6]

type designation structure: nnnn.nnnn X nn (where n = 0 to 9 and X may be "." "K"), all models rated DC 10 - 27 V [TSME], DC 10 - 28 V [TSME6]; 1.8 A, for use with external power supplies [AC 100 - 240 V; 47 - 63 Hz; 0.8...0.4 A] or [AC 100 - 240 V; 50/60 Hz; 1.4 A], Pollution Degree 2, Overvoltage Category II.



Certificate: 70002782
Project: 80038448

Master Contract: 190420
Date Issued: 2020-04-14

APPLICABLE REQUIREMENTS

- | | |
|------------------------------|--|
| CAN/CSA-C22.2 No. 61010-1-12 | - Safety Requirements for Electrical Equipment for Measurement, Controls, and Laboratory Use, Part 1: General Requirements |
| UL 61010-1 (Third Edition) | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |

Conditions of Acceptability:

- (1) The equipment is supplied with an approved external power supply, by Cincon Electronics, type TRG30R120V or Mean Well GST120A15 with an approved power supply cord set or power supply cord with plug that is acceptable to the authorities in the country where the equipment is to be used.