ROHDE&SCHWARZ





This is to certify that:

| Equipment Type | Stock No. | Designation |
|----------------|----------------|--|
| FSW8 | 1331.5003n08 | Signal and spectrum analyzer, 2 Hz to 8 GHz |
| FSW13 | 1331.5003n13 | Signal and spectrum analyzer, 2 Hz to 13.6 GHz |
| FSW26 | 1331.5003n26 | Signal and spectrum analyzer, 2 Hz to 26.5 GHz |
| FSW43 | 1331.5003n43 | Signal and spectrum analyzer, 2 Hz to 43.5 GHz |
| FSW50 | 1331.5003n50 | Signal and spectrum analyzer, 2 Hz to 50 GHz |
| FSW67 | 1331.5003n67 | Signal and spectrum analyzer, 2 Hz to 67 GHz |
| FSW85 | 1331.5003n85 | Signal and spectrum analyzer, 2 Hz to 85 GHz |
| | n = "." or "K" | |

complies with the provisions of the Directive of the Council of the European Union on the approximation of the laws of the Member States.

EMC Directive (2014/30/EU)

Harmonized standards applied: EN 55011:2016 + A1:2017 + A11:2020

EN 61326-1:2013

Low Voltage Directive (2014/35/EU)

Harmonized standards applied: EN 61010-1:2010+A1:2019+A1:2019/AC:2019

Restriction of the Use of Certain Hazardous Substances Directive (2011/65/EU)

Harmonized standards applied: EN IEC 63000:2018

The immunity test requirements for equipment intended to be used in industrial electromagnetic environment (acc. to EN 61326-1, Table 2) have been used as a basis.

The limits of radio interference for Group I, Class A equipment (acc. to EN 55011) have been used as a basis.

This declaration of conformity is issued under the sole responsibility of the manufacturer.

CE

ROHDE & SCHWARZ GmbH & Co. KG Mühldorfstr. 15, D-81671 Munich V. ZET G