ROHDE&SCHWARZ





This is to certify that:

Equipment Type	Stock No.	Designation
FSV3004	1330.5000n04	SIGNAL ANALYZER 4GHz
FSVA3004	1330.5000n05	SIGNAL ANALYZER 4GHZ
FSV3007	1330.5000n07	SIGNAL ANALYZER 7.5GHZ
FSVA3007	1330.5000n08	SIGNAL ANALYZER 7.5GHZ
FSV3013	1330.5000n13	SIGNAL ANALYZER 13.6GHz
FSVA3013	1330.5000n14	SIGNAL ANALYZER 13.6GHZ
FSV3030	1330.5000n30	SIGNAL ANALYZER 30GHz
FSVA3030	1330.5000n31	SIGNAL ANALYZER 30GHZ
FSV3044	1330.5000n43	SIGNAL ANALYZER 44GHz
FSVA3044	1330.5000n44	SIGNAL ANALYZER 44GHZ
FSV3050	1330.5000n50	SIGNAL ANALYZER 50GHz
FSVA3050	1330.5000n51	SIGNAL ANALYZER 50GHz
	n = "." or "K"	

complies with the provisions of the Directive of the Council of the European Union on the approximation of the laws of the Member States.

EMC Directive (2014/30/EU)

Harmonized standards applied: EN 55011:2016 + A1:2017 + A11:2020

EN 61326-1:2013

Low Voltage Directive (2014/35/EU)

Harmonized standards applied: EN 61010-1:2010+A1:2019+A1:2019/AC:2019

Restriction of the Use of Certain Hazardous Substances Directive (2011/65/EU)

Harmonized standards applied: EN IEC 63000:2018

The immunity test requirements for equipment intended to be used in industrial electromagnetic environment (acc. to EN 61326-1, Table 2) have been used as a basis.

The limits of radio interference for Group I, Class A equipment (acc. to EN 55011) have been used as a basis.

This declaration of conformity is issued under the sole responsibility of the manufacturer.

ROHDE & SCHWARZ GmbH & Co. KG Mühldorfstr. 15, D-81671 Munich

₩ QM CI