



Certificate of Compliance

Certificate: 70120054

Master Contract: 190420

Project: 70120054

Date Issued: 2017-06-07

Issued to: Rohde & Schwarz Gmbh & Co. KG
Postfach 80 14 69
Muehldorfstrasse 15
Muenchen/Bayern, 81614
GERMANY
Attention: Max Weissenbacher

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: *Carsten Kuehler*
Carsten Kuehler

PRODUCTS

CLASS - C363106 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS - C363186 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE-Certified to US Standards

Power Meter: models 1424.7005X02 [NRX]; 1430.0509X02 [NRT2];

Type designation structure: nnnn.nnnn X nn (where n = 0 to 9 and X may be "." or "K"), rated AC 100...240 V; 50...60 / 400 Hz; 1.7 A, pollution degree 2, installation category II

APPLICABLE REQUIREMENTS

- | | |
|------------------------------|----------------------------------------------------------------------------------------------------------------------------|
| CAN/CSA-C22.2 No. 61010-1-12 | - Safety Requirements for Electrical Equipment for Measurement, Controls, and Laboratory Use, Part 1: General Requirements |
| UL 61010-1 (Third Edition) | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |

CONDITIONS OF ACCEPTABILITY

- (1) The mains supply cord set or power supply cord with plug (< 3 m) provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.