



Declaration of Conformity

This is to certify that:

Equipment Type	Stock No.	Designation
ZNB MKII - ZNB4 MKII	1334.3330n22	VECT.NETW.ANALYZER 2 TOR
ZNB MKII - ZNB4 MKII	1334.3330n24	VECT.NETW.ANALYZER 4 TOR
ZNB MKII - ZNB8 MKII	1334.3330n42	VECT.NETW.ANALYZER 2 TOR
ZNB MKII - ZNB8 MKII	1334.3330n44	VECT.NETW.ANALYZER 4 TOR
ZNB MKII - ZNB20 MKII	1334.3330n62	VECT.NETW.ANALYZER 2 TOR
ZNB MKII - ZNB26 MKII	1334.3330n63	VECT.NETW.ANALYZER 2 TOR
ZNB MKII - ZNB20 MKII	1334.3330n64	VECT.NETW.ANALYZER 4 TOR
ZNB MKII - ZNB26 MKII	1334.3330n65	VECT.NETW.ANALYZER 4 TOR
ZNB MKII - ZNB40 MKII	1334.3330n72	VECT.NETW.ANALYZER 2 TOR
ZNB MKII - ZNB43 MKII	1334.3330n92	VECT.NETW.ANALYZ. 2T 2.92
ZNB MKII - ZNB43 MKII	1334.3330n93	VECT.NETW.ANALYZ. 2T 2.4
ZNB MKII - ZNB43 MKII	1334.3330n94	VECT.NETW.ANALYZ. 4T 2.92
ZNB MKII - ZNB43 MKII	1334.3330n95	VECT.NETW.ANALYZ. 4T 2.4

n= "." or "K"

complies with the provisions of the Directive of the Council of the European Union on the approximation of the laws of the Member States.

**EMC Directive (2014/30/EU)**

Harmonized standards applied: EN 55011:2016 + A1:2017 + A11:2020  
 EN 61000-3-2:2014  
 EN 61000-3-3:2013  
 EN 61326-1:2013  
 EN 61326-2-1:2013

**Low Voltage Directive (2014/35/EU)**

Harmonized standards applied: EN 61010-1:2010+A1:2019+A1:2019/AC:2019

**Restriction of the Use of Certain Hazardous Substances Directive (2011/65/EU)**

Harmonized standards applied: EN IEC 63000:2018

The immunity test requirements for equipment intended to be used in industrial electromagnetic environment (acc. to EN 61326-1, Table 2) have been used as a basis.  
 The limits of radio interference for Group I, Class A equipment (acc. to EN 55011) have been used as a basis.

ROHDE & SCHWARZ GmbH & Co. KG  
 Mühldorfstr. 15, D-81671 Munich

*i.v. [Signature]*