



Certificate of Compliance

Certificate: 70214983

Master Contract: 190420

Project: 80151106

Date Issued: 2023-03-28

Issued To: Rohde & Schwarz Gmbh & Co. KG Abt. 6FRK
Postfach 80 14 69
Muehldorfstrasse 15
Muenchen, Bavaria, 81614
Germany

Attention: Max Weissenbacher

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: *Natascha Heger*
Natascha Heger



PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Certified to US Standards

Radio Communication Tester, Models 1201.0002X20 [CMP200], 1201.0002X18 [CMP180], 1201.0002X36 [PVT360A]

Type designation structure: nnnn.nnnn X nn (where n = 0 to 9 and X may be "." or "K"),

Input: AC 100-240 V, 50-60 Hz, max. 7.3 - 4.6 A or 8.9 - 3.7 A (depending on SMPSU), detachable cord set,

Equipment class I (PE connected), Pollution Degree 2, Overvoltage Category II, continuous operation, environmental conditions: 5°C to 45°C, altitude max. 2000m, 80% rel. humidity.



Certificate: 70214983
Project: 80151106

Master Contract: 190420
Date Issued: 2023-03-28

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12,
UPD1:2015, UPD2:2016, AMD1:2018

- Safety Requirements for Electrical Equipment for Measurement, Controls, and Laboratory Use, Part 1: General Requirements

UL 61010-1 Third Edition (2012),
AMD1:2018

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

Notes:

Products certified under Class C363106, C363186 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca

