



Certificate of Compliance

Certificate: 80004420

Master Contract: 190420

Project: 80082364

Date Issued: 2021-06-29

Issued To: Rohde & Schwarz Gmbh & Co. KG Abt. 6FRK
Postfach 80 14 69
Muehldorfstrasse 15
Muenchen, Bavaria, 81614
Germany

Attention: Max Weissenbacher

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: *Christoph Choudhuri*
Christoph Choudhuri



PRODUCTS

CLASS – 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS – 3631 86 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Certified to US Standards

Radio Communication Tester CMX, Models 1201.0002X70 [CMX500], Type designation structure: nnnn.nnnn X nn (where n = 0 to 9 and X may be "." or "K"), all models are cord connected, rated AC 100-240 V, 50-60 Hz, max. 8.9 – 3.7 A (if equipped with PSU800 2 times) or 16...12.5 A (if equipped with PSX2800), all models are cord connected(detachable cord set), Equipment Class I (PE connected), Pollution Degree 2, Overvoltage Category II, continuous operation, operating altitude max. 2000 m, operating temperature +5°C to +45°C, max. 80 % r.H.(non-condensing).



Certificate: 80004420
Project: 80082364

Master Contract: 190420
Date Issued: 2021-06-29

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12,
UPD1:2015, UPD2:2016, AMD1:2018

- Safety Requirements for Electrical Equipment for Measurement, Controls, and Laboratory Use, Part 1: General Requirements

UL 61010-1 Third Edition (2012),
AMD1:2018

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

Conditions of Acceptability:

- (1) The mains supply cord set or power supply cord with plug (< 3 m) provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.