

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product

Electrical equipment for measurement and control use

Name and address of the applicant

Rohde & Schwarz GmbH & Co. KG  
Mühlendorfstraße 15, 81671 München  
GERMANY

Name and address of the manufacturer

Rohde & Schwarz GmbH & Co. KG  
Mühlendorfstraße 15, 81671 München  
GERMANY

Name and address of the factory

ROHDE & SCHWARZ MESSGERÄTEBAU GmbH  
Rohde-und-Schwarz-Strasse 1, 87700 Memmingen  
GERMANY*Note: When more than one factory, please report on page 2*

Ratings and principal characteristics

 Additional Information on page 2Input 1: AC 100...240 V; 50...60 Hz; 8.9...3.7 A  
Input 2: AC 100...240 V; 50...60 Hz; 8.9...3.7 A  
or AC 100...240 V; 50...60 Hz; 16...12.5 A

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

CTF Stage 1

Model / Type Ref.

CMX500 [1201.0002X70]

Additional information (if necessary may also be reported on page 2)

Type designation structure:

nnnn.nnnnXnn (where n = 0 to 9 and X may be "." or "K")

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010  
IEC 61010-1:2010/AMD1:2016

As shown in the Test Report Ref. No. which forms part of this Certificate

273073-TL4-1

This CB Test Certificate is issued by the National Certification Body

**VDE** Prüf- und Zertifizierungsinstitut GmbH  
**VDE** Testing and Certification Institute  
Zertifizierungsstelle und internat. Angelegenheiten  
Certification Body and internat. Affairs



G. Heine

Date: 2021-07-02

Signature: