



Certificate of Compliance

Certificate: 80021036

Master Contract: 190420

Project: 80115445

Date Issued: 2022-03-08

Issued To: Rohde & Schwarz Gmbh & Co. KG Abt. 6FRK
Postfach 80 14 69
Muehldorfstrasse 15
Muenchen, Bavaria, 81614
Germany

Attention: Max Weissenbacher

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Andreas Kohler

Issued by: Andreas Kohler



PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Certified to US Standards

Vector Signal Generator SMCV100B, Models 1432.7000X02 [SMCV100B], Type designation structure: nnnn.nnnn X nn (where n = 0 to 9 and X may be "." or "K"), all models are cord connected (detachable cord-set), rated AC 100-240 V, 50-60 Hz, max. 3.6 - 1.5 A, Pollution Degree 2, Overvoltage Category II, 4600 m Altitude, operating ambient +5 to +45°C, rel. humidity max. 90% (non-condensing).

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12,
UPD1:2015, UPD2:2016, AMD1:2018

- Safety Requirements for Electrical Equipment for Measurement, Controls, and Laboratory Use, Part 1: General Requirements

UL 61010-1 3rd ed (2012) + AMD1
(2018)

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Certificate: 80021036
Project: 80115445

Master Contract: 190420
Date Issued: 2022-03-08

Notes:

Products certified under Class C363106, C363186 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca

