

R&S®SMM-K131

LoRa®

User Manual



1179206002
Version 05

ROHDE & SCHWARZ
Make ideas real



This document describes the following software option:

- R&S®SMM-K131 LoRa® (1441.1760.xx)

This manual describes firmware version FW 5.30.047.xx and later of the R&S®SMM100A.

© 2023 Rohde & Schwarz

Muehldorfstr. 15, 81671 Muenchen, Germany

Phone: +49 89 41 29 - 0

Email: info@rohde-schwarz.com

Internet: www.rohde-schwarz.com

Subject to change – data without tolerance limits is not binding.

R&S® is a registered trademark of Rohde & Schwarz GmbH & Co. KG.

LoRa® Mark and LoRa Logo are registered trademarks of Semtech Corporation.

LoRaWAN® Mark and LoRaWAN Logo are registered trademarks of LoRa Alliance, Inc.

All other trademarks are the properties of their respective owners.

1179.2060.02 | Version 05 | R&S®SMM-K131

The following abbreviations are used throughout this manual: R&S®SMM100A is abbreviated as R&S SMM, R&S®WinIQSIM2 is abbreviated as R&S WinIQSIM2; the license types 02/03/07/11/13/16/12 are abbreviated as xx.

Contents

1	Welcome to the LoRa option.....	5
1.1	Accessing the LoRa dialog.....	5
1.2	What's new.....	5
1.3	Documentation overview.....	6
1.3.1	Getting started manual.....	6
1.3.2	User manuals and help.....	6
1.3.3	Service manual.....	6
1.3.4	Instrument security procedures.....	7
1.3.5	Printed safety instructions.....	7
1.3.6	Data sheets and brochures.....	7
1.3.7	Release notes and open source acknowledgment (OSA).....	7
1.3.8	Application notes, application cards, white papers, etc.....	7
1.3.9	Videos.....	7
1.4	Scope.....	8
1.5	Notes on screenshots.....	8
2	About the LoRa option.....	9
2.1	Required options.....	9
2.2	About LoRa.....	9
2.3	About LoRaWAN.....	11
3	LoRa configuration and settings.....	14
3.1	General settings.....	14
3.2	Frame configuration settings.....	17
3.3	Impairments settings.....	20
4	Signal generation control.....	23
4.1	Trigger settings.....	23
4.2	Marker settings.....	28
4.3	Clock settings.....	30
4.4	Local and global connectors settings.....	31
5	Generating LoRa test signals.....	32
5.1	Generating RX sensitivity test signals.....	32

5.2 Generating blocking test signals.....	33
6 Remote-control commands.....	35
6.1 Programming examples.....	35
6.2 General commands.....	38
6.3 Frame configuration commands.....	41
6.4 Impairments commands.....	46
6.5 Trigger commands.....	48
6.6 Marker commands.....	53
6.7 Clock commands.....	55
Glossary: Abbreviations and definitions.....	56
Glossary: Specifications and references.....	57
List of commands.....	58
Index.....	60

1 Welcome to the LoRa option

The R&S SMM-K131 is a firmware application that adds functionality to generate signals in accordance with LoRa modulation.

The R&S SMM-K131 features

- Generation of LoRa up-/downlink signals
- Generation of waveforms for LoRa up-/downlink signals
- Frame configuration including modulation, coding and data configuration
- Impairing the output signal:
 - Adding symbol timing error and frequency offset
 - Applying a frequency drift to the carrier frequency of the output signal

This user manual contains a description of the functionality that the application provides, including remote control operation.

All functions not discussed in this manual are the same as in the base unit and are described in the R&S SMM100A user manual. The latest version is available at:

www.rohde-schwarz.com/manual/SMM100A

Installation

You can find detailed installation instructions in the delivery of the option or in the R&S SMM100A service manual.

1.1 Accessing the LoRa dialog

To open the dialog with LoRa settings

- ▶ In the block diagram of the R&S SMM100A, select "Baseband > LoRa".
A dialog box opens that displays the provided general settings.

The signal generation is not started immediately. To start signal generation with the default settings, select "State > On".

1.2 What's new

This manual describes firmware version FW 5.30.047.xx and later of the R&S®SMM100A.

Compared to the previous version, it provides the new features listed below:

- Time-based triggering, see "[Time Based Trigger](#)" on page 25 and "[Trigger Time](#)" on page 25.

- Editorial changes

1.3 Documentation overview

This section provides an overview of the R&S SMM100A user documentation. Unless specified otherwise, you find the documents at:

www.rohde-schwarz.com/manual/smm100a

1.3.1 Getting started manual

Introduces the R&S SMM100A and describes how to set up and start working with the product. Includes basic operations, typical measurement examples, and general information, e.g. safety instructions, etc. A printed version is delivered with the instrument.

1.3.2 User manuals and help

Separate manuals for the base unit and the software options are provided for download:

- Base unit manual
Contains the description of all instrument modes and functions. It also provides an introduction to remote control, a complete description of the remote control commands with programming examples, and information on maintenance, instrument interfaces and error messages. Includes the contents of the getting started manual.
- Software option manual
Contains the description of the specific functions of an option. Basic information on operating the R&S SMM100A is not included.

The contents of the user manuals are available as help in the R&S SMM100A. The help offers quick, context-sensitive access to the complete information for the base unit and the software options.

All user manuals are also available for download or for immediate display on the Internet.

1.3.3 Service manual

Describes the performance test for checking compliance with rated specifications, firmware update, troubleshooting, adjustments, installing options and maintenance.

The service manual is available for registered users on the global Rohde & Schwarz information system (GLORIS):

<https://gloris.rohde-schwarz.com>

1.3.4 Instrument security procedures

Deals with security issues when working with the R&S SMM100A in secure areas. It is available for download on the internet.

1.3.5 Printed safety instructions

Provides safety information in many languages. The printed document is delivered with the product.

1.3.6 Data sheets and brochures

The data sheet contains the technical specifications of the R&S SMM100A. It also lists the options and their order numbers and optional accessories.

The brochure provides an overview of the instrument and deals with the specific characteristics.

See www.rohde-schwarz.com/brochure-datasheet/smm100a

1.3.7 Release notes and open source acknowledgment (OSA)

The release notes list new features, improvements and known issues of the current firmware version, and describe the firmware installation.

The software makes use of several valuable open source software packages. An open-source acknowledgment document provides verbatim license texts of the used open source software.

See www.rohde-schwarz.com/firmware/smm100a

1.3.8 Application notes, application cards, white papers, etc.

These documents deal with special applications or background information on particular topics.

See www.rohde-schwarz.com/application/smm100a

1.3.9 Videos

Find various videos on Rohde & Schwarz products and test and measurement topics on YouTube: <https://www.youtube.com/@RohdeundSchwarz>



On the menu bar, search for your product to find related videos.

HOME VIDEOS SHORTS PLAYLISTS COMMUNITY CHANNELS ABOUT



Figure 1-1: Product search on YouTube

1.4 Scope



Tasks (in manual or remote operation) that are also performed in the base unit in the same way are not described here.

In particular, it includes:

- Managing settings and data lists, like saving and loading settings, creating and accessing data lists, or accessing files in a particular directory.
- Information on regular trigger, marker and clock signals and filter settings, if appropriate.
- General instrument configuration, such as checking the system configuration, configuring networks and remote operation
- Using the common status registers

For a description of such tasks, see the R&S SMM100A user manual.

1.5 Notes on screenshots

When describing the functions of the product, we use sample screenshots. These screenshots are meant to illustrate as many as possible of the provided functions and possible interdependencies between parameters. The shown values may not represent realistic usage scenarios.

The screenshots usually show a fully equipped product, that is: with all options installed. Thus, some functions shown in the screenshots may not be available in your particular product configuration.

2 About the LoRa option

2.1 Required options

The basic equipment layout for generating LoRa signals includes the options:

- Baseband Generator(R&S SMM-B9)
- Baseband real-time extension (R&S SMM-K520)
- Option LoRa (R&S SMM-K131)
- Frequency option (e.g. R&S SMM-B1006)

You can generate signals via play-back of waveform files at the signal generator. To create the waveform file using R&S WinIQSIM2, you do not need a specific option.

To play back the waveform file at the signal generator, you have two options:

- Install the R&S WinIQSIM2 option of the digital standard, e.g. R&S SMM-K255 for playing LTE waveforms
- If supported, install the real-time option of the digital standard, e.g. R&S SMM-K55 for playing LTE waveforms

For more information, see data sheet.

2.2 About LoRa


The term LoRa refers to a "Long Range" end-to-end communication technology based on a **PHY** with a proprietary **modulation** scheme. The standard is owned by Semtech Corporation, Camarillo California, USA.

Modulation


LoRa modulation is based on a proprietary **CSS** modulation scheme. The modulation scheme is used to encode data onto swept frequency chirps, i.e. rapid changes of the modulation frequency.

A chirp (compressed high intensity radar pulse) is a signal whose frequency changes at a fixed rate (chirp rate). Each modulation symbol is represented as a rapid change in the frequency of the chirp.

The spreading factor represents the number of encoded bits per symbol and thus the change in frequency per unit time.



The higher the spreading factor, the lower the change in frequency per unit time.



Chirp rate R_C , signal bandwidth Δf , spreading factor SF, sample rate R_S , symbol duration T_S , bit rate R_B and coding rate CR are linked as in the equations below. Furthermore, typical values are given for SF, Δf and CR.

Symbol rate: $R_S = 1 / T_S = \Delta f / 2^{SF} \mid 125 \text{ kHz} \leq \Delta f \leq 500 \text{ kHz};$

Chirp rate: $R_C = R_S * 2^{SF} = \Delta f \mid 7 \leq SF \leq 12;$

Bit rate: $R_B = SF * \Delta f / 2^{SF} * 4 / (4 + CR) \mid 1 \leq CR \leq 4$

A coding rate CR = 0 corresponds to no coding.

For related settings, see "[Modulation and Coding Configuration](#)" on page 18.

Channel allocation

LoRa communication channels can be freely attributed by the network following the spectrum allocation rules defined by the regional radio regulation authorities. In Europe, the allocation rules are defined by [ETSI](#) in the specifications [ETSI Specification EN 300 220-1](#) and [CEPT/ERC Recommendation 70-03](#). In North America, the allocation rules are defined by [FCC](#) in the [FCC Specification 47 Part 15.247](#).

For detailed info about regional LoRa performance requirements, see [LoRaWAN Regional Parameters Specification](#).


Table 2-1: Example: LoRaWAN regional parameters for Europe and North America

Parameter	Europe (ETSI)	North America (FCC)
Frequency band	863 MHz to 869 MHz	902 MHz to 928 MHz
Number of channels	Up- and downlink: 10	Uplink: 64 x 125 kHz and 8 x 500 kHz Downlink: 8 x 500 kHz

Parameter	Europe (ETSI)	North America (FCC)
Channel bandwidth	Uplink: 125 kHz and 250 kHz Downlink 125 kHz	Uplink: 125 kHz and 500 kHz Downlink: 500 kHz
TX power	Uplink: 14 dBm (20 dBm allowed) Downlink: 14 dBm	Uplink: 20 dBm (30 dBm allowed) Downlink: 27 dBm
Spreading factor	Uplink: 7 to 12	Uplink: 7 to 12
Data rate	250 bit/s to 50000 bit/s	980 bit/s to 21900 bit/s
Link budget	Uplink: 155 dB Downlink: 155 dB	Uplink: 154 dB Downlink: 157 dB

For related settings, see [Chapter 3.1, "General settings"](#), on page 14.

Message and frame structure



For related settings, see [Chapter 3.2, "Frame configuration settings"](#), on page 17.

2.3 About LoRaWAN

The LoRa Alliance, Inc. specifies a wide area network (WAN) stack for long range communication as shown in [Figure 2-1 \(1MA295\)](#). The LoRa modulation physical layer enables the long-range communication link. The LoRa MAC and application layers affect battery lifetime of the end-device, network capacity, quality of service and security.




Figure 2-1: LoRaWAN stack

Network architecture

A LoRaWAN architecture has a star-shaped structure as shown in [Figure 2-2 \(1MA295\)](#). End-devices exchange data with the network server via gateways. Also, the network server is connected to application servers, on which typically IoT applications run.




Figure 2-2: LoRaWAN network architecture

For more information, see the [LoRaWAN Specification](#).

Device communication classes

There are three classes specified for LoRaWAN compliant devices:

- **Class A**: Bi-directional end-devices (mandatory support)
 - Bi-directional communication between LoRa network server and receiver. A scheduled uplink transmission slot (Transmit) is followed by two downlink receive slots (Rx 1 and Rx 2).




Figure 2-3: Class A communication

- **Class B:** Bi-directional end-devices with scheduled receive slots (optional support)
Besides Class A communication, more downlink receive slots (Rx) are available due to a time synchronized periodic beacon signal (Beacon) from network gateway.




Figure 2-4: Class B communication

- **Class C:** Bi-directional end-devices with maximal receive slots (optional support)
Besides Class A and B communication, there are continuously open receive slots except during transmission.




Figure 2-5: Class C communication

Typically, power consumption increases and latency decreases from Class A to Class C communication.

Key features and applications

Devices compliant with the LoRa technology and operating in a LoRaWAN offer the following key features:

- **Long range:** outdoor coverage of up to 30 miles/48.3 km (line of sight)
- **Low power consumption:** battery lifetime of up to 20 years
- **Low cost:** low-cost end-devices and open software

The key features meet the requirements for IoT applications in rural areas.

3 LoRa configuration and settings

Access:

- ▶ Select "Baseband > LoRa".

The remote commands required to define these settings are described in [Chapter 6, "Remote-control commands"](#), on page 35.


Settings:

● General settings	14
● Frame configuration settings	17
● Impairments settings	20

3.1 General settings

Access:

- ▶ Select "Baseband > LoRa".



The tab comprises the standard general settings.

Settings:

State	15
Set to Default	15
Save/Recall	15
Generate Waveform File	15
Bandwidth	16
Idle Interval	16

Sequence Length.....	16
Oversampling.....	16
Sample Rate Variation.....	16

State

Activates the standard and deactivates all the other digital standards and digital modulation modes in the same path.

Remote command:

[\[:SOURce<hw>\]:BB:LORA:STATE](#) on page 41

Set to Default

Calls the default settings. The values of the main parameters are listed in the following table.

Parameter	Value
"State"	Not affected by the "Set to Default"
"Bandwidth"	125 kHz
"Idle Interval"	100.0 us
"Sequence Length"	1 frame
"Oversampling"	4
"Sample Rate Variation"	500.000 000 kHz

Remote command:

[\[:SOURce<hw>\]:BB:LORA:PRESet](#) on page 39

Save/Recall

Accesses the "Save/Recall" dialog, that is the standard instrument function for saving and recalling the complete dialog-related settings in a file. The provided navigation possibilities in the dialog are self-explanatory.

The settings are saved in a file with predefined extension. You can define the filename and the directory, in that you want to save the file.

See also, chapter "File and Data Management" in the R&S SMM100A user manual.

Remote command:

[\[:SOURce<hw>\]:BB:LORA:SETTing:CATalog](#) on page 39

[\[:SOURce<hw>\]:BB:LORA:SETTing:STORe](#) on page 40

[\[:SOURce<hw>\]:BB:LORA:SETTing:LOAD](#) on page 39

[\[:SOURce<hw>\]:BB:LORA:SETTing:DELeTe](#) on page 39

Generate Waveform File

With enabled signal generation, triggers the instrument to save the current settings of an arbitrary waveform signal in a waveform file with predefined extension *.wv. You can define the filename and the directory, in that you want to save the file.

Using the ARB modulation source, you can play back waveform files and/or process the file to generate multi-carrier or multi-segment signals.

Remote command:

[:SOURce<hw>] :BB:LORA:WAVEform:CREate on page 41

Bandwidth

Sets the channel bandwidth.

The sample rate and FFT size are calculated internally and updated automatically.

Remote command:

[:SOURce<hw>] :BB:LORA:BWIDth on page 38

Idle Interval

Sets the time of the interval separating two frames.

Remote command:

[:SOURce<hw>] :BB:LORA:IINTerval on page 38

Sequence Length

Sets the sequence length of the signal in number of frames. The signal is calculated in advance and output in the arbitrary waveform generator.

Remote command:

[:SOURce<hw>] :BB:LORA:SLENgth on page 40

Oversampling

Sets the oversampling factor of the generated waveform. The ARB generator of the R&S SMM100A requires low oversampling factors and still provides excellent signal quality in terms of EVM and ACP.

A reduced sample rate saves significantly the amount of memory or allows an increased signal cycle time, and vice versa.

Remote command:

[:SOURce<hw>] :BB:LORA:OSAMpling on page 40

Sample Rate Variation

Sets the sample rate of the signal. A variation of this parameter affects the ARB clock rate; all other signal parameters remain unchanged.

When changing values of the affecting parameters, the sample rate is reset according to the equations below:

- **Symbol Timing Error** disabled:
Sample rate = **Bandwidth** * **Oversampling**
- **Impairments** enabled:
Sample rate = (abs(**Freq. drift deviation**) + abs(**Freq. offset**) + (bandwidth * over-sampling) / 2) * 2


Remote command:

[:SOURce<hw>] :BB:LORA:SRATE:VARIation on page 40

3.2 Frame configuration settings

Access:

- ▶ Select "Baseband > LoRa > Frame Configuration".



This tab comprises the general and data settings to configure the frame structure.

Settings

Payload Reduced Coding Mode	17
Unmodulated Preamble Length	17
Sync Mode	18
Modulation and Coding Configuration	18
└ Coding Rate	18
└ Spreading Factor	18
└ Encoder Active	18
└ Interleaver Active	18
Data	18
└ Data Length	19
└ Data Source	19
└ Payload CRC	20
└ Header Configuration	20
└ Header Active	20
└ Compressed Mode	20
└ Burst Mode	20
└ Reserved Bit	20

Payload Reduced Coding Mode

Activates the payload reduced coding mode.

Remote command:

[:SOURce<hw>] :BB:LORA:FCONfiguration:PRCMode:STATE on page 44

Unmodulated Preamble Length

Sets the unmodulated preamble length.

Remote command:

[**:SOURce<hw>]:BB:LORA:FCONfiguration:UPLength** on page 45

Sync Mode

Sets the synchronization mode of the preamble.

Public A preamble with a public sync word is generated.

Private A preamble with a private sync word is generated.

Remote command:

[**:SOURce<hw>]:BB:LORA:FCONfiguration:SMODE** on page 45

Modulation and Coding Configuration

Configures the modulation and coding information in the frame configuration.

Coding Rate ← Modulation and Coding Configuration

Sets the coding rate. The coding rate R_{Coding} is calculated as follows:

$$R_{Coding} = 4 / (4 + \text{"Coding Rate"})$$

"Coding Rate = 0" corresponds to no coding, i.e. $R_{Coding} = 1$.

Remote command:

[**:SOURce<hw>]:BB:LORA:FCONfiguration:CRATE** on page 42

Spreading Factor ← Modulation and Coding Configuration

Sets the spreading factor for the modulation.

Remote command:

[**:SOURce<hw>]:BB:LORA:FCONfiguration:SFACtor** on page 45

Encoder Active ← Modulation and Coding Configuration

Activates encoding of the modulation symbols.

Remote command:

[**:SOURce<hw>]:BB:LORA:FCONfiguration:EACTive:STATE** on page 43

Interleaver Active ← Modulation and Coding Configuration

Activates the interleaver in the frame.

Remote command:


[**:SOURce<hw>]:BB:LORA:FCONfiguration:IACTive:STATE** on page 44

Data

Access:

Select "Baseband > LoRa> Frame Configuration > Data".

Configures header and payload data settings.



Data Length ← Data

Sets the data length of the payload in the frame.

Remote command:

[\[:SOURce<hw>\]:BB:LORA:FCONfiguration:DLENgth](#) on page 43

Data Source ← Data

Selects the data source for the payload.

The following standard data sources are available:

- "All 0, All 1"
An internally generated sequence containing 0 data or 1 data.
- "PNxx"
An internally generated pseudo-random noise sequence.
- "Pattern"
An internally generated sequence according to a bit pattern.
Use the "Pattern" box to define the bit pattern.
- "Data List>Select DList"
A binary data from a data list, internally or externally generated.
Select "Select DList" to access the standard "Select List" dialog.
 - Select the "Select Data List > navigate to the list file *.dm_iqd > Select" to select an existing data list.
 - Use the "New" and "Edit" functions to create internally new data list or to edit an existing one.
 - Use the standard "File Manager" function to transfer external data lists to the instrument.

See also:

- Section "Modulation Data" in the R&S SMM100A user manual.
- Section "File and Data Management" in the R&S SMM100A user manual.
- Section "Data List Editor" in the R&S SMM100A user manual

Remote command:

[:SOURce<hw>] :BB:LORA:FConfiguration:DATA on page 42
[:SOURce<hw>] :BB:LORA:FConfiguration:DATA:DPAattern on page 43
[:SOURce<hw>] :BB:LORA:FConfiguration:DATA:DSELection on page 43

Payload CRC ← Data

Activates a cyclic redundancy check (CRC) of the payload.

The "Payload CRC" bits are appended to the payload of the frame.

Remote command:

[:SOURce<hw>] :BB:LORA:FConfiguration:PCRC:STATE on page 44

Header Configuration ← Data

Configures the header information in the payload.

Header Active ← Header Configuration ← Data

Activates the header data in the frame.

Remote command:

[:SOURce<hw>] :BB:LORA:FConfiguration:HACTive:STATE on page 44

Compressed Mode ← Header Configuration ← Data

Activates the compressed mode of the header data in the frame.

Remote command:

[:SOURce<hw>] :BB:LORA:FConfiguration:CMODE:STATE on page 42

Burst Mode ← Header Configuration ← Data

Activates the burst mode of header data in the frame.

Remote command:

[:SOURce<hw>] :BB:LORA:FConfiguration:BMODE:STATE on page 41

Reserved Bit ← Header Configuration ← Data

Enables the reserved bit in the frame header.


Remote command:

[:SOURce<hw>] :BB:LORA:FConfiguration:RBIT:STATE on page 45

3.3 Impairments settings

Access:

- ▶ Select "Baseband > LoRa > Impairments".



This tab comprises the impairments settings for the signal.

Settings

Impairments	21
Symbol Timing Error	21
Frequency Offset	21
Frequency Drift Configuration	21
└ State	21
└ Deviation	22
└ Type	22
└ Rate	22

Impairments

Activates the impairments settings.

Impairments change the signal to simulate a non-ideal transmitter.

Remote command:

[\[:SOURce<hw>\]:BB:LORA:IMPaIrments:STATE](#) on page 47

Symbol Timing Error

Sets the symbol timing error.

The set error corresponds to applying a deviation to the transmitter symbol clock.

Remote command:

[\[:SOURce<hw>\]:BB:LORA:IMPaIrments:STERror](#) on page 47

Frequency Offset

Sets the carrier frequency offset.

Remote command:

[\[:SOURce<hw>\]:BB:LORA:IMPaIrments:FOFFset](#) on page 47

Frequency Drift Configuration

Configures carrier frequency drift settings of the signal.

State ← Frequency Drift Configuration

Activates carrier frequency drift settings.

Remote command:

[**:SOURce<hw>]:BB:LORA:IMPaIrments:FDRift:STATE** on page 46

Deviation ← Frequency Drift Configuration

Sets the maximum deviation of the carrier signal from the center frequency during the frequency drift procedure.

Remote command:

[**:SOURce<hw>]:BB:LORA:IMPaIrments:FDDeviAtion** on page 46

Type ← Frequency Drift Configuration

Sets the type of frequency drift, i.e. how the carrier signal drifts around the center frequency.

Linear Linear frequency drift.

Sine Sine frequency drift.

Remote command:

[**:SOURce<hw>]:BB:LORA:IMPaIrments:FDTyPe** on page 46

Rate ← Frequency Drift Configuration

Sets the rate of the carrier frequency drift.

A rate of, e.g., 300 Hz implies, that the carrier signal drifts 300 times per second around the center frequency.

Remote command:

[**:SOURce<hw>]:BB:LORA:IMPaIrments:FDRate** on page 46

4 Signal generation control

This section lists settings provided for configuring the baseband filter, for defining the signal generation start and for generating signals necessary for synchronization with other instruments.


Settings:

• Trigger settings.....	23
• Marker settings.....	28
• Clock settings.....	30
• Local and global connectors settings.....	31

4.1 Trigger settings

Access:

- ▶ Select "Baseband" > "LoRa" > "Trigger In".



This tab provides settings to select and configure the trigger, like trigger source, trigger mode and trigger delays, and to arm or trigger an internal trigger manually. The header of the tab displays the status of the trigger signal and trigger mode. As in the tabs "Marker" and "Clock", this tab provides also access to the settings of the related connectors.

Routing and activating a trigger signal

1. Define the effect of a trigger event and the trigger signal source.
 - a) Select "Trigger In" > "Mode".
 - b) Select "Trigger In" > "Source".
2. For external trigger signals, define the connector for signal input. See [Chapter 4.4, "Local and global connectors settings", on page 31](#). You can map trigger signals to one or more User x or T/M connectors.
Local and global connectors settings allow you to configure the signal mapping, the polarity, the trigger threshold and the input impedance of the input connectors.
3. Activate baseband signal generation. In the block diagram, set "Baseband" > "On". The R&S SMM100A starts baseband signal generation after the configured trigger event.

About baseband trigger signals

This section focuses on the available settings.

For information on how these settings affect the signal, refer to section "Basics on ..." in the R&S SMM100A user manual.

Settings:

Mode	24
Signal Duration Unit	24
Signal Duration	25
Running/Stopped	25
Time Based Trigger	25
Trigger Time	25
Arm	26
Execute Trigger	26
Source	26
Sync. Output to External Trigger/Sync. Output to Trigger	26
External Inhibit/Trigger Inhibit	27
External Delay/Trigger Delay	27

Mode

Selects trigger mode, i.e. determines the effect of a trigger event on the signal generation.

- "Auto"
The signal is generated continuously.
- "Retrigger"
The signal is generated continuously. A trigger event (internal or external) causes a restart.
- "Armed Auto"
The signal is generated only when a trigger event occurs. Then the signal is generated continuously.
An "Arm" stops the signal generation. A subsequent trigger event (internal or external) causes a restart.
- "Armed Retrigger"
The signal is generated only when a trigger event occurs. Then the signal is generated continuously. Every subsequent trigger event causes a restart.
An "Arm" stops signal generation. A subsequent trigger event (internal or external) causes a restart.
- "Single"
The signal is generated only when a trigger event occurs. Then the signal is generated once to the length specified at "Signal Duration".
Every subsequent trigger event (internal or external) causes a restart.

Remote command:

[\[:SOURce<hw>\] :BB:LORA \[:TRIGger\] :SEQUence](#) on page 48

Signal Duration Unit

Defines the unit for describing the length of the signal sequence to be output in the "Single" trigger mode.

Remote command:

[:SOURce<hw>] :BB:LORA:TRIGger:SLUNit on page 51

Signal Duration

Requires trigger "Mode" > "Single".

Enters the length of the trigger signal sequence.

Use this parameter, for example, for the following applications:

- To output the trigger signal partly.
- To output a predefined sequence of the trigger signal.

Remote command:

[:SOURce<hw>] :BB:LORA:TRIGger:SLENgth on page 51

Running/Stopped

With enabled modulation, displays the status of signal generation for all trigger modes.

- "Running"
The signal is generated; a trigger was (internally or externally) initiated in triggered mode.
- "Stopped"
The signal is not generated and the instrument waits for a trigger event.

Remote command:

[:SOURce<hw>] :BB:LORA:TRIGger:RMODE on page 51

Time Based Trigger

Requires trigger "Mode" > "Armed Auto"/"Single".

Activates time-based triggering with a fixed time reference.

The R&S SMM100A triggers signal generation when its operating system time ("Current Time") matches a specified time trigger ("Trigger Time"). As trigger source, you can use an internal trigger or an external global trigger.

How to: Chapter "Time-based triggering" in the R&S SMM100A user manual.

Remote command:

[:SOURce<hw>] :BB:LORA:TRIGger:TIME[:STATe] on page 52

Trigger Time

Requires trigger "Mode" > "Armed Auto"/"Single".

Sets date and time for a time-based trigger signal.

Set a trigger time that is later than the "Current Time". The current time is the operating system time of the R&S SMM100A. If you set an earlier trigger time than the current time, time-based triggering is not possible.

How to: Chapter "Time-based triggering" in the R&S SMM100A user manual.

"Date" Sets the date of the time-based trigger in format YYYY-MM-DD.

Remote command:

[:SOURce<hw>] :BB:LORA:TRIGger:TIME:DATE on page 51

"Time" Sets the time of the time-based trigger in format hh:mm:ss.
 Remote command:
`[:SOURce<hw>] :BB:LORA:TRIGger:TIME:TIME` on page 52

Arm

Stops the signal generation until subsequent trigger event occurs.

Remote command:

`[:SOURce<hw>] :BB:LORA:TRIGger:ARM:EXECute` on page 50

Execute Trigger

For internal trigger source, executes trigger manually.

Remote command:

`[:SOURce<hw>] :BB:LORA:TRIGger:EXECute` on page 50

Source

The following sources of the trigger signal are available:

- "Internal"
The trigger event is executed manually by the "Execute Trigger".
- "External Global Trigger"
The trigger event is the active edge of an external trigger signal provided and configured at the User x connectors.
- "Baseband Sync In"
In primary-secondary instrument mode, secondary instruments are triggered by the active edge of the synchronization signal.

How to: ["Routing and activating a trigger signal"](#) on page 23


Remote command:

`[:SOURce<hw>] :BB:LORA:TRIGger:SOURce` on page 48

Sync. Output to External Trigger/Sync. Output to Trigger

Enables signal output synchronous to the trigger event.


- "On"
Corresponds to the default state of this parameter.
The signal calculation starts simultaneously with the trigger event. Because of the processing time of the instrument, the first samples are cut off and no signal is output. After elapsing of the internal processing time, the output signal is synchronous to the trigger event.



- "Off"

The signal output begins after elapsing of the processing time. Signal output starts with sample 0. The complete signal is output.

This mode is recommended for triggering of short signal sequences. Short sequences are sequences with signal duration comparable with the processing time of the instrument.



Remote command:

`[:SOURce<hw>] :BB:LORA:TRIGger[:EXTernal] :SYNChronize:OUTPut`
on page 49

External Inhibit/Trigger Inhibit

Applies for external trigger signal.

Sets the duration with that any following trigger event is suppressed. In "Retrigger" mode, for example, a new trigger event does not cause a restart of the signal generation until the specified inhibit duration does not expire.

For more information, see chapter "Basics" in the R&S SMM100A user manual.

Remote command:

`[:SOURce<hw>] :BB:LORA:TRIGger[:EXTernal] :INHibit` on page 49

External Delay/Trigger Delay

Delays the trigger event of the signal from:

- The external trigger source

Use this setting to:

- Synchronize the instrument with the device under test (DUT) or other external devices

For more information, see chapter "Basics on ..." in the R&S SMM100A user manual.


Remote command:

`[:SOURce<hw>] :BB:LORA:TRIGger[:EXTernal] :DELay` on page 49

4.2 Marker settings

Access:

- Select "Baseband" > "LoRa" > "Marker".



This tab provides settings to select and configure the marker output signal including marker mode and marker delay.

Routing and activating a marker signal

1. To define the signal shape of an individual marker signal "x", select "Marker" > "Marker x" > "Mode".
2. Optionally, define the connector for signal output. See [Chapter 4.4, "Local and global connectors settings"](#), on page 31.
You can map marker signals to one or more User x or T/M connectors.
3. Activate baseband signal generation. In the block diagram, set "Baseband" > "On".
The R&S SMM100A adds the marker signal to the baseband signal. Also, R&S SMM100A outputs this signal at the configured User x connector.

About marker output signals

This section focuses on the available settings.

For information on how these settings affect the signal, refer to section "Basics on ..." in the R&S SMM100A user manual.

Settings:

Mode.....	28
Delay.....	29

Mode

Marker configuration for up to 3 markers. The settings are used to select the marker mode defining the shape and periodicity of the markers. The contents of the dialog change with the selected marker mode.

How to: ["Routing and activating a marker signal"](#) on page 28

"Restart" A marker signal is generated at the start of each ARB sequence.

"Frame" A frame clock with the frame period specified under "Period" is generated on the output connector. The marker signal is generated after every specified number of frames.

"Pulse" A regular marker signal is generated. The pulse frequency is defined by entering a divider. The frequency is derived by dividing the sample rate by the divider. The input box for the divider opens when "Pulse" is selected, and the resulting pulse frequency is displayed below it. The maximum pulse frequency is half the symbol rate.

Remote command:

[\[:SOURce<hw>\]:BB:LORA:TRIGger:OUTPut<ch>:PULSe:DIVider](#) on page 54

[\[:SOURce<hw>\]:BB:LORA:TRIGger:OUTPut<ch>:PULSe:FREQuency?](#) on page 54

"Pattern" A marker signal that is defined by a bit pattern is generated. The pattern has a maximum length of 64 bits and is defined in an input field which opens when pattern is selected.

Remote command:

[\[:SOURce<hw>\]:BB:LORA:TRIGger:OUTPut<ch>:PATtern](#) on page 54

"On/Off Ratio" A regular marker signal that is defined by an On/Off ratio is generated. A period lasts one On and Off cycle.



Remote command:

[\[:SOURce<hw>\]:BB:LORA:TRIGger:OUTPut<ch>:ONTIme](#) on page 54

[\[:SOURce<hw>\]:BB:LORA:TRIGger:OUTPut<ch>:OFFTime](#) on page 54

Remote command:

[\[:SOURce<hw>\]:BB:LORA:TRIGger:OUTPut<ch>:MODE](#) on page 53

Delay

Delays the marker signal at the marker output relative to the signal generation start.

Variation of the parameter "Marker x" > "Delay" causes signal recalculation.


Remote command:

[\[:SOURce<hw>\]:BB:LORA:TRIGger:OUTPut<ch>:DELy](#) on page 53

4.3 Clock settings

Access:

- ▶ Select "Baseband" > "LoRa" > "Clock".



This tab provides settings to select and configure the clock signal, like the clock source and clock mode.

Defining the clock

1. Select "Clock" > "Source" to define the source of clock signal.
2. For external clock signals, define the connector for signal input. See [Chapter 4.4, "Local and global connectors settings", on page 31](#).
You can map clock signals to one or more User x or T/M connectors.
Local and global connectors settings allow you to configure the signal mapping, the polarity, the trigger threshold and the input impedance of the input connectors.
3. Activate baseband signal generation. In the block diagram, set "Baseband" > "On".
The R&S SMM100A starts baseband signal generation with a symbol rate that equals the clock rate.

About clock signals

This section focuses on the available settings.

For information on how these settings affect the signal, refer to section "Basics on ..." in the R&S SMM100A user manual.

Settings:

Clock Source	30
Clock Mode	31

Clock Source

Selects the clock source.

- "Internal"
The instrument uses its internal clock reference.

How to: ["Defining the clock" on page 30](#)

Remote command:

[:SOURce<hw>] :BB:LORA:CLOCK:SOURce on page 55

Clock Mode

Sets the type of externally supplied clock.

Remote command:

[:SOURce<hw>] :BB:LORA:CLOCK:MODE on page 55

4.4 Local and global connectors settings

Accesses a dialog to configure local connectors or global connectors.

The button is available in the following dialogs or tabs:

- "Trigger / Marker / Clock" dialog that is accessible via the "TMC" block in the block diagram.
- "Trigger In", "Marker" and "Clock" tabs that are accessible via the "Baseband" block in the block diagram.



See also chapter "Local and global connectors settings" in the user manual.

5 Generating LoRa test signals

The section describes how to set up and configure the R&S SMM100A for LoRa signal generation for simple receiver tests.

See also:

- Application note [1MA295](#), for a comprehensive overview on LoRa device characterization, including Tx measurements.
- [Chapter 6.1, "Programming examples"](#), on page 35, for remote command configuration examples.


5.1 Generating RX sensitivity test signals

Rx sensitivity test setup

The R&S SMM100A generates a LoRa test signal. The LoRa DUT receives and demodulates the test signal. The LoRa Test Tool, installed on an external PC, analyzes the demodulation data in terms of [PER](#) analysis.

To perform an Rx sensitivity test

1. Preset the R&S SMM100A.
2. Select "Baseband > LoRa".
3. Configure "General" settings of the LoRa test signal, e.g. the "Bandwidth" compatible with the DUT.
4. Select "Frame Configuration > General" to configure modulation and coding parameters, e.g. "Coding Rate" and "Spreading Factor".
5. Select "Frame Configuration > Data" to configure payload, e.g. "Data Length" and "Data Source".
6. Set "General > State > On" to activate LoRa baseband modulation.
7. In the block diagram, set "I/Q Mod > On" and "RF > On" to activate I/Q modulation and signal generation.
8. On the "Status Bar", set "Frequency" and "Level".



9. Configure the DUT for reception of the LoRa test signal generated at the R&S SMM100A.
 10. Gradually reduce the level of the LoRa test signal.
 11. Monitor the **PER** at the DUT, e.g. at an external PC with an analysis software (LoRa Test Tool).
- The Rx sensitivity level corresponds to a specified PER value, e.g. "Level > -137 dBm" with PER = 1 %. The level depends on the spreading factor.

5.2 Generating blocking test signals

Blocking test setup

The R&S SMM100A generates a LoRa test signal, the wanted signal. A second signal generator, e.g. an R&S SGS, generates an interfering **CW** signal, the unwanted signal. Both signals are combined via an RF power combiner. The LoRa DUT receives and demodulates the combined signal. The LoRa Test Tool, installed on external PC, analyzes the demodulation data in terms of **PER** analysis.

To perform an adjacent channel blocking test

1. To generate the wanted signal, configure the R&S SMM100A as in "To perform an **Rx sensitivity test**" on page 32.

2. To generate the adjacent channel blocking signal at the second signal generator, e.g. the R&S SGS, proceed as follows:
 - a) Preset the R&S SGS.
 - b) Set the frequency as specified.
The frequency value is typically a few hundred kilohertz higher/lower than the frequency of the wanted signal.
 - c) Set the level as specified.
The level values for uplink and downlink can differ.

3. At the R&S SMM100A, increase the level of the wanted signal.

4. Monitor the **PER** at the DUT, e.g. with an analysis software (LoRa Test Tool).

The adjacent channel blocking level is the wanted signal level, that corresponds to a specified PER value, typically PER < 1 %.

To perform an on-channel blocking test

1. To generate the wanted signal, configure the R&S SMM100A as in "[To perform an Rx sensitivity test](#)" on page 32.
2. To generate the on-channel blocking signal at the second signal generator, e.g. the R&S SGS, proceed as follows:
 - a) Preset the R&S SGS.
 - b) Set the frequency, that is the same frequency as the wanted signal ([step 8](#)).
 - c) Set the level as specified.
The level values for uplink and downlink can differ.

3. At the R&S SMM100A, increase the level of the wanted signal.

4. Monitor the **PER** at the DUT, e.g. with an analysis software (LoRa Test Tool).

The on-channel blocking level is the wanted signal level, that corresponds to a specified PER value, typically PER < 1 %.

See also:

- [R&S SGS100A user manual](#)
- [R&S SGMA-GUI Software](#) software for control of the R&S SGS

6 Remote-control commands

The following commands are required to generate signals with the LoRa option in a remote environment. We assume that the R&S SMM100A has already been set up for remote operation in a network as described in the R&S SMM100A documentation. A knowledge about the remote control operation and the SCPI command syntax are assumed.



Conventions used in SCPI command descriptions

For a description of the conventions used in the remote command descriptions, see section "Remote-Control Commands" in the R&S SMM100A user manual.

Common suffixes

The following common suffixes are used in the remote commands:

Suffix	Value range	Description
SOURce<hw>	1	Available baseband signals
OUTPut<ch>	1 to 3	Available markers

The following commands specific to the R&S SMM100A-K131 option are described here:

● Programming examples	35
● General commands	38
● Frame configuration commands	41
● Impairments commands	46
● Trigger commands	48
● Marker commands	53
● Clock commands	55

6.1 Programming examples

The corresponding sections of the same title provide simple programming examples for the R&S SMM100A. The purpose of the examples is to present **all** commands for a given task. In real applications, one would rather reduce the examples to an appropriate subset of commands.

The programming examples have been tested with a software tool which provides an environment for the development and execution of remote tests. To keep the examples as simple as possible, only the "clean" SCPI syntax elements are reported. Non-executable command lines (for example comments) start with two // characters.

At the beginning of the most remote control program, an instrument (p)reset is recommended to set the R&S SMM100A to a definite state. The commands *RST and SYSTem:PRESet are equivalent for this purpose. *CLS also resets the status registers and clears the output buffer.

Example: Storing current configuration

```
SOURCE1:BB:LORA:SETTING:STORe "/var/user/my_settings"  
*RST  
SOURCE1:BB:LORA:SETTING:CATalog?  
// my_lora, lora  
SOURCE1:BB:LORA:SETTING:LOAD "/var/user/lora"  
SOURCE1:BB:LORA:STATE 1  
SOURCE1:BB:LORA:SETTING:DEL "my_lora"  
SOURCE1:BB:LORA:WAVEform:CREAtE "/var/user/my_lora_wv"
```

Example: Generating a LoRa signal

```
SOURCE1:BB:LORA:PRESet  
// Configure LoRa general settings: Bandwidth, idle interval time, sequence  
// length and sampling parameters  
SOURCE1:BB:LORA:BWIDth BW125  
SOURCE1:BB:LORA:IINTerval 0.0001  
SOURCE1:BB:LORA:SLENgth 1  
SOURCE1:BB:LORA:OSAMpling 4  
SOURCE1:BB:LORA:SRATE:VARiation 500000  
SOURCE1:BB:LORA:STATE 1  
// Configure the carrier signal, e.g. for ETSI compliant devices.  
SOURCE1:FREQuency:CW 868500000  
SOURCE1:POWER:POWeR 14  
OUTPUT1:STATE 1
```

Example: Configuring LoRa frame parameters

```
SOURCE1:BB:LORA:PRESet  
// Configure preamble and synchronization parameters.  
SOURCE1:BB:LORA:FConfiguration:PRCMode:STATE 0  
SOURCE1:BB:LORA:FConfiguration:SMODE PUBL  
SOURCE1:BB:LORA:FConfiguration:UPLength 8  
  
// Configure modulation and coding parameters.  
SOURCE1:BB:LORA:FConfiguration:CRATe CR1  
SOURCE1:BB:LORA:FConfiguration:EACTive:STATE 1  
SOURCE1:BB:LORA:FConfiguration:SFACtor SF7  
SOURCE1:BB:LORA:FConfiguration:IACTive:STATE 1  
  
// Configure data, header and payload parameters.  
SOURCE1:BB:LORA:FConfiguration:DLENgth 16  
SOURCE1:BB:LORA:FConfiguration:PCRC:STATE 1  
SOURCE1:BB:LORA:FConfiguration:DATA PN9  
SOURCE1:BB:LORA:FConfiguration:HACTive:STATE 1  
SOURCE1:BB:LORA:FConfiguration:BMODe:STATE 0  
SOURCE1:BB:LORA:FConfiguration:CMODe:STATE 0  
SOURCE1:BB:LORA:FConfiguration:RBIT:STATE 0
```

Example: Configuring impairments to LoRa modulation

```
SOURCE1:BB:LORA:PRESet
// Configure symbol timing error and frequency offset.
SOURCE1:BB:LORA:IMPAirments:STERror 0
SOURCE1:BB:LORA:IMPAirments:FOFFset 0
// Configure frequency drift parameters.
SOURCE1:BB:LORA:IMPAirments:FDTYpe LIN
SOURCE1:BB:LORA:IMPAirments:FDDeviAtion 0
SOURCE1:BB:LORA:IMPAirments:FDRate 300
SOURCE1:BB:LORA:IMPAirments:FDRift:STATe 1
// Activate the impairment settings.
SOURCE1:BB:LORA:IMPAirments:STATe 1
```

Example: Trigger configuration

```
SOURCE1:BB:LORA:TRIGger:SOURce INTernal
SOURCE1:BB:LORA:TRIGger:SEQUence AREtrigger
SOURCE1:BB:LORA:STAT ON
SOURCE1:BB:LORA:TRIGger:EXECute
SOURCE1:BB:LORA:TRIGger:ARM:EXECute
SOURCE1:BB:LORA:TRIGger:RMODE?
// stopped
SOURCE1:BB:LORA:TRIGger:EXECute
SOURCE1:BB:LORA:TRIGger:RMODE?

// run
SOURCE1:BB:LORA:TRIGger:SEQUence SING
SOURCE1:BB:LORA:TRIGger:SLUNit SAMP
SOURCE1:BB:LORA:TRIGger:SLENGth 25777
SOURCE1:BB:LORA:TRIGger:SEQUence ARET
SOURCE1:BB:LORA:TRIGger:SOURce EGT1
SOURCE1:BB:LORA:TRIGger:EXTernal:SYNChronize:OUTPut 1
SOURCE1:BB:LORA:TRIGger:EXTernal:INHibit 100
SOURCE1:BB:LORA:TRIGger:EXTernal:DELay 10
```

Example: Configure and enable standard marker signals

```
// Configure marker 1 settings.
SOURCE:BB:LORA:TRIGger:OUTPut1:MODE RAT
SOURCE:BB:LORA:TRIGger:OUTPut1:ONTime 10
SOURCE:BB:LORA:TRIGger:OUTPut1:OFFTime 20
// Configure marker 2 settings.
SOURCE:BB:LORA:TRIGger:OUTPut2:MODE?
// RESTart
SOURCE:BB:LORA:TRIGger:OUTPut2:DELay 16384
// Configure marker 3 settings.
SOURCE:BB:LORA:TRIGger:OUTPut3:MODE PATT
SOURCE:BB:LORA:TRIGger:OUTPut3:PATTern #H2,2
```

Example: Clock settings

```
SOURCE1:BB:LORA:CLOCK:SOURce ELCL
SOURCE1:BB:LORA:CLOCK:MODE SAMP
CLOCK:INPUT:FREQuency?
// 1000000
```

6.2 General commands

[:SOURce<hw>]:BB:LORA:BWIDth.....	38
[:SOURce<hw>]:BB:LORA:IINTerval.....	38
[:SOURce<hw>]:BB:LORA:PRESet.....	39
[:SOURce<hw>]:BB:LORA:SETting:CATalog.....	39
[:SOURce<hw>]:BB:LORA:SETting:DElete.....	39
[:SOURce<hw>]:BB:LORA:SETting:LOAD.....	39
[:SOURce<hw>]:BB:LORA:SETting:STORe.....	40
[:SOURce<hw>]:BB:LORA:SLENgth.....	40
[:SOURce<hw>]:BB:LORA:OSAMpling.....	40
[:SOURce<hw>]:BB:LORA:SRATe:VARiation.....	40
[:SOURce<hw>]:BB:LORA:STATE.....	41
[:SOURce<hw>]:BB:LORA:WAVEform:CREate.....	41

[:SOURce<hw>]:BB:LORA:BWIDth <Bw>

Sets the channel bandwidth.

Parameters:

<Bw>	BW7 BW10 BW15 BW20 BW31 BW41 BW62 BW125 BW250 BW500
	*RST: BW125

Example: See [Example "Generating a LoRa signal"](#) on page 36.

Manual operation: See ["Bandwidth"](#) on page 16

[:SOURce<hw>]:BB:LORA:IINTerval <IInterval>

Sets the time of the interval separating two frames.

Parameters:

<IInterval>	float
	Range: 0 to 1
	Increment: 0.1E-6
	*RST: 100E-6
	Default unit: s

Example: See [Example "Generating a LoRa signal"](#) on page 36.

Manual operation: See ["Idle Interval"](#) on page 16

[:SOURce<hw>]:BB:LORA:PRESet****

Sets the parameters of the digital standard to their default values (*_{RST} values specified for the commands).

Not affected is the state set with the command **SOURCE<hw> : BB : LORA : STATE**.

Example: See [Example "Generating a LoRa signal"](#) on page 36.

Usage: Event

Manual operation: See ["Set to Default"](#) on page 15

[:SOURce<hw>]:BB:LORA:SETTING:CATalog <Catalog>****

Queries the files with settings in the default directory. Listed are files with the file extension *.lora.

Parameters:

<Catalog> <filename1>,<filename2>,...

Returns a string of filenames separated by commas.

Example: See [Example "Storing current configuration"](#) on page 36.

Manual operation: See ["Save/Recall"](#) on page 15

[:SOURce<hw>]:BB:LORA:SETTING:DELetE <Filename>****

Deletes the selected file from the default or the specified directory. Deleted are files with extension *.lora.

Parameters:

<Filename> "<filename>"

Filename or complete file path; file extension can be omitted

Example: See [Example "Storing current configuration"](#) on page 36.

Manual operation: See ["Save/Recall"](#) on page 15

[:SOURce<hw>]:BB:LORA:SETTING:LOAD <Filename>****

Loads the selected file from the default or the specified directory. Loaded are files with extension *.lora.

Parameters:

<Filename> "<filename>"

Filename or complete file path; file extension can be omitted

Example: See [Example "Storing current configuration"](#) on page 36.

Manual operation: See ["Save/Recall"](#) on page 15

[:SOURce<hw>]:BB:LORA:SETTING:STORe <Filename>****

Saves the current settings into the selected file; the file extension (*.lora) is assigned automatically.

Setting parameters:

<Filename> "<filename>"
 Filename or complete file path

Example: See [Example "Storing current configuration" on page 36](#).

Usage: Setting only

Manual operation: See ["Save/Recall"](#) on page 15

[:SOURce<hw>]:BB:LORA:SLength <SLength>****

Sets the sequence length of the signal in number of frames. The signal is calculated in advance and output in the arbitrary waveform generator.

Parameters:

<SLength> integer
 Range: 1 to dynamic
 *RST: 1

Example: See [Example "Generating a LoRa signal" on page 36](#).

Manual operation: See ["Sequence Length"](#) on page 16

[:SOURce<hw>]:BB:LORA:OSAMpling <OSampling>****

Sets the oversampling factor of the generated waveform.

A reduced sample rate saves significantly the amount of memory or allows an increased signal cycle time, and vice versa.

Parameters:

<OSampling> integer
 Range: 1 to 32
 *RST: 4

Example: See [Example "Generating a LoRa signal" on page 36](#).

Manual operation: See ["Oversampling"](#) on page 16

[:SOURce<hw>]:BB:LORA:SRATE:VARiation <Variation>****

Sets the sample rate of the signal.

A variation of this parameter affects the ARB clock rate; all other signal parameters remain unchanged. When changing values of the affecting parameters, the sample rate is reset.

Parameters:

<Variation> float
 Range: 4E2 to 20E6
 Increment: 1E-3
 *RST: 1E6

Example: See [Example "Generating a LoRa signal" on page 36](#).

Manual operation: See ["Sample Rate Variation"](#) on page 16

[:SOURce<hw>]:BB:LORA:STATe <State>

Activates the standard and deactivates all the other digital standards and digital modulation modes in the same path.

Parameters:

<State> 1 | ON | 0 | OFF
 *RST: 0

Example: See [Example "Generating a LoRa signal" on page 36](#).

Manual operation: See ["State"](#) on page 15

[:SOURce<hw>]:BB:LORA:WAVeform:CREate <Filename>

Saves the current settings as an ARB signal in a waveform file (*.wv).

Setting parameters:

<Filename> string
 Filename or complete file path; file extension is assigned automatically

Example: See [Example "Storing current configuration" on page 36](#).

Usage: Setting only

Manual operation: See ["Generate Waveform File"](#) on page 15

6.3 Frame configuration commands

[:SOURce<hw>]:BB:LORA:FCONfiguration:BMODe:STATe <State>

Activates the burst mode of header data in the frame.

Parameters:

<State> 1 | ON | 0 | OFF
 *RST: 0

Example: See [Example "Configuring LoRa frame parameters" on page 36](#).

Manual operation: See ["Burst Mode"](#) on page 20

[:SOURce<hw>]:BB:LORA:FCONfiguration:CMODE:STATE <State>

Activates the compressed mode of the header data in the frame.

Parameters:

<State> 1 | ON | 0 | OFF
 *RST: 0

Example: See [Example"Configuring LoRa frame parameters" on page 36](#).

Manual operation: See "[Compressed Mode](#)" on page 20

[:SOURce<hw>]:BB:LORA:FCONfiguration:CRATE <CRate>

Sets the coding rate.

Parameters:

<CRate> CR0 | CR1 | CR2 | CR3 | CR4
 $CR<x> = 0 \text{ to } 4$
 The coding rate R_{Coding} is calculated as follows:
 $R_{Coding} = 4 / (4 + CR<x>)$
 "CR0" corresponds to no coding, i.e. $R_{Coding} = 1$.
 *RST: CR1

Example: See [Example"Configuring LoRa frame parameters" on page 36](#).

Manual operation: See "[Coding Rate](#)" on page 18

[:SOURce<hw>]:BB:LORA:FCONfiguration:DATA <Data>

Sets the data source for the payload data in a LoRa frame.

Parameters:

<Data> ZERO | ONE | PATTern | PN9 | PN11 | PN15 | PN16 | PN20 |
 PN21 | PN23 | DLIST

PNxx

The pseudo-random sequence generator is used as the data source. There is a choice of different lengths of random sequence.

DLIST

A data list is used. The data list is selected with the aid of command :SOURce:BB:LORA:DATA DLIST.

ALL0 | ALL1

Internal 0 or 1 data is used.

PATTern

Internal data is used. The bit pattern for the data is defined with the aid of command :SOURce:BB:LORA:DATA PATTern.

*RST: PN9

Example: See [Example"Configuring LoRa frame parameters" on page 36](#).

Manual operation: See "Data Source" on page 19

[:SOURce<hw>]:BB:LORA:FCONfiguration:DATA:DPATtern <DPattern>, <BitCount>

Sets the data pattern, if the `data source` PATT is selected.

Parameters:

<DPattern>	numeric
	*RST: #H0
<BitCount>	integer
	Range: 1 to 64
	*RST: 1

Example: See Example "Configuring LoRa frame parameters" on page 36.

Manual operation: See "Data Source" on page 19

[:SOURce<hw>]:BB:LORA:FCONfiguration:DATA:DSELection <DSelection>

Selects an existing data list file from the default directory or from the specific directory.

The data list is only used, if the `data source` DLIS is selected.

Parameters:

<DSelection>	string
	Filename incl. file extension or complete file path

Example: See Example "Configuring LoRa frame parameters" on page 36.

Manual operation: See "Data Source" on page 19

[:SOURce<hw>]:BB:LORA:FCONfiguration:DLENgth <DLength>

Sets the data length of the payload in the frame.

Parameters:

<DLength>	integer
	Range: 1 to 255
	*RST: 16

Example: See Example "Configuring LoRa frame parameters" on page 36.

Manual operation: See "Data Length" on page 19

[:SOURce<hw>]:BB:LORA:FCONfiguration:EACTive:STATE <State>

Activates encoding of the modulation symbols.

Parameters:

<State>	1 ON 0 OFF
	*RST: 1

Example: See [Example "Configuring LoRa frame parameters" on page 36](#).

Manual operation: See ["Encoder Active"](#) on page 18

[:SOURce<hw>]:BB:LORA:FCONfiguration:HACTive:STATe <State>

Activates the header data in the frame.

Parameters:

<State> 1 | ON | 0 | OFF
*RST: 1

Example: See [Example "Configuring LoRa frame parameters" on page 36](#).

Manual operation: See ["Header Active"](#) on page 20

[:SOURce<hw>]:BB:LORA:FCONfiguration:IACTive:STATe <State>

Activates the interleaver in the frame.

Parameters:

<State> 1 | ON | 0 | OFF
*RST: 1

Example: See [Example "Configuring LoRa frame parameters" on page 36](#).

Manual operation: See ["Interleaver Active"](#) on page 18

[:SOURce<hw>]:BB:LORA:FCONfiguration:PCRC:STATe <State>

Activates a cyclic redundancy check (CRC) of the payload.

Parameters:

<State> 1 | ON | 0 | OFF
*RST: 1

Example: See [Example "Configuring LoRa frame parameters" on page 36](#).

Manual operation: See ["Payload CRC"](#) on page 20

[:SOURce<hw>]:BB:LORA:FCONfiguration:PRCMode:STATe <State>

Activates the payload reduced coding mode.

Parameters:

<State> 1 | ON | 0 | OFF
*RST: 0

Example: See [Example "Configuring LoRa frame parameters" on page 36](#).

Manual operation: See ["Payload Reduced Coding Mode"](#) on page 17

[:SOURce<hw>]:BB:LORA:FCONfiguration:RBIT:STATe <State>

Enables the reserved bit in the frame header.

Parameters:

<State>	1 ON 0 OFF
	*RST: 0

Example: See [Example"Configuring LoRa frame parameters" on page 36](#).

Manual operation: See ["Reserved Bit"](#) on page 20

[:SOURce<hw>]:BB:LORA:FCONfiguration:SMode <SMode>

Sets the synchronization mode of the preamble.

Parameters:

<SMode>	PRIvate PUBLic
	PRIvate
	A preamble with a public sync word is generated.
	PUBLic
	A preamble with a private sync word is generated.
	*RST: PUBLic

Example: See [Example"Configuring LoRa frame parameters" on page 36](#).

Manual operation: See ["Sync Mode"](#) on page 18

[:SOURce<hw>]:BB:LORA:FCONfiguration:SFACtor <Sf>

Sets the spreading factor for the modulation.

Parameters:

<Sf>	SF6 SF7 SF8 SF9 SF10 SF11 SF12
	*RST: SF7

Example: See [Example"Configuring LoRa frame parameters" on page 36](#).

Manual operation: See ["Spreading Factor"](#) on page 18

[:SOURce<hw>]:BB:LORA:FCONfiguration:UPLength <PLength>

Sets the unmodulated preamble length.

Parameters:

<PLength>	integer
	Range: 6 to 8
	*RST: 8

Example: See [Example"Configuring LoRa frame parameters" on page 36](#).

Manual operation: See ["Unmodulated Preamble Length"](#) on page 17

6.4 Impairments commands

[**:SOURce<hw>]:BB:LORA:IMPaIrments:FDDeViAtion <FDDeViAtion>**

Sets the frequency deviation of the frequency drift.

Parameters:

<FDDeViAtion>	integer
	Range: -200E3 to 200E3
	*RST: 0

Example: See [Example "Configuring impairments to LoRa modulation"](#) on page 37.

Manual operation: See "[Deviation](#)" on page 22

[**:SOURce<hw>]:BB:LORA:IMPaIrments:FDRate <FDRate>**

Sets the rate of the carrier frequency drift.

Parameters:

<FDRate>	integer
	Range: 160 to 1.6E3
	*RST: 300

Example: See [Example "Configuring impairments to LoRa modulation"](#) on page 37.

Manual operation: See "[Rate](#)" on page 22

[**:SOURce<hw>]:BB:LORA:IMPaIrments:FDRift:STATe <State>**

Activates frequency drift settings in the payload.

Parameters:

<State>	1 ON 0 OFF
	*RST: 1

Example: See [Example "Configuring impairments to LoRa modulation"](#) on page 37.

Manual operation: See "[State](#)" on page 21

[**:SOURce<hw>]:BB:LORA:IMPaIrments:FDTYpe <FDType>**

Sets the frequency drift type.

Parameters:

<FDType>	LINear SINE
	LINear Generation of frequency drift is set linear.

SINE

Generation of frequency drift is set sinusoid.

*RST: LINear

Example: See [Example "Configuring impairments to LoRa modulation"](#) on page 37.

Manual operation: See "[Type](#)" on page 22

[:SOURce<hw>]:BB:LORA:IMPaIrments::FOFFset <FOffset>

Sets the frequency offset.

Parameters:

<FOffset>	integer Range: -200E3 to 200E3 *RST: 0
--	--

Example: See [Example "Configuring impairments to LoRa modulation"](#) on page 37.

Manual operation: See "[Frequency Offset](#)" on page 21

[:SOURce<hw>]:BB:LORA:IMPaIrments::STATe <State>

Activates impairments settings in the payload.

Parameters:

<State>	1 ON 0 OFF *RST: 0
--	-----------------------------

Example: See [Example "Configuring impairments to LoRa modulation"](#) on page 37.

Manual operation: See "[Impairments](#)" on page 21

[:SOURce<hw>]:BB:LORA:IMPaIrments::STERror <STERrror>

Sets symbol timing error.

Parameters:

<STERrror>	integer Range: -300 to 300 *RST: 0
---	--

Example: See [Example "Configuring impairments to LoRa modulation"](#) on page 37.

Manual operation: See "[Symbol Timing Error](#)" on page 21

6.5 Trigger commands

[:SOURce<hw>]:BB:LORA[:TRIGger]:SEQuence.....	48
[:SOURce<hw>]:BB:LORA:TRIGger:SOURce.....	48
[:SOURce<hw>]:BB:LORA:TRIGger[:EXTernal]:SYNChronize:OUTPut.....	49
[:SOURce<hw>]:BB:LORA:TRIGger[:EXTernal]:DELay.....	49
[:SOURce<hw>]:BB:LORA:TRIGger[:EXTernal]:INHibit.....	49
[:SOURce<hw>]:BB:LORA:TRIGger:OBASEband:DELay.....	49
[:SOURce<hw>]:BB:LORA:TRIGger:OBASEband:DELay.....	50
[:SOURce<hw>]:BB:LORA:TRIGger:OBASEband:INHibit.....	50
[:SOURce<hw>]:BB:LORA:TRIGger:ARM:EXECute.....	50
[:SOURce<hw>]:BB:LORA:TRIGger:EXECute.....	50
[:SOURce<hw>]:BB:LORA:TRIGger:RMODE.....	51
[:SOURce<hw>]:BB:LORA:TRIGger:SLENGTH.....	51
[:SOURce<hw>]:BB:LORA:TRIGger:SLUnit.....	51
[:SOURce<hw>]:BB:LORA:TRIGger:TIME:DATE.....	51
[:SOURce<hw>]:BB:LORA:TRIGger:TIME:TIME.....	52
[:SOURce<hw>]:BB:LORA:TRIGger:TIME[:STATe].....	52

[:SOURce<hw>]:BB:LORA[:TRIGger]:SEQuence <Sequence>

Sets the trigger mode.

Parameters:

<Sequence> AUTO | RETRigger | AAUTo | ARETrigger | SINGLE
 *RST: AUTO

Example: See [Example "Trigger configuration"](#) on page 37.

Manual operation: See ["Mode"](#) on page 24

[:SOURce<hw>]:BB:LORA:TRIGger:SOURce <Source>

Selects the trigger signal source and determines the way the triggering is executed.

Provided are:

- Internal triggering by a command (`INTERNAL`)
- External trigger signal via one of the local or global connectors
 - EGT1 | EGT2: External global trigger
 - EGC1 | EGC2: External global clock
- In primary-secondary instrument mode, the external baseband synchronization signal (`BBSY`)
- OBASEband | BEXTernal | EXTernal: Setting only
Provided only for backward compatibility with other Rohde & Schwarz signal generators.

The R&S SMM100A accepts these values and maps them automatically as follows:

`EXTernal` = EGT1, `BEXTernal` = EGT2, `OBASEband` = `INTA`

Parameters:

<Source> INTernal|EGT1|EGT2|EGC1|EGC2|EXTernal|BBSY
 *RST: INTernal

Example: See [Example"Trigger configuration"](#) on page 37.

Manual operation: See "[Source](#)" on page 26

[:SOURce<hw>]:BB:LORA:TRIGger[:EXTernal]:SYNChronize:OUTPut <Output>

Enables signal output synchronous to the trigger event.

Parameters:

<Output> 1 | ON | 0 | OFF
 *RST: 1

Example: See [Example"Trigger configuration"](#) on page 37.

Manual operation: See "[Sync. Output to External Trigger/Sync. Output to Trigger](#)" on page 26

[:SOURce<hw>]:BB:LORA:TRIGger[:EXTernal]:DELay <Delay>

Sets the trigger delay.

Parameters:

<Delay> float
 Range: 0 to 2147483647
 Increment: 0.01
 *RST: 0

Example: See [Example"Trigger configuration"](#) on page 37.

Manual operation: See "[External Delay/Trigger Delay](#)" on page 27

[:SOURce<hw>]:BB:LORA:TRIGger[:EXTernal]:INHibit <Inhibit>

Specifies the duration by which a restart is inhibited.

Parameters:

<Inhibit> integer
 Range: 0 to dynamic
 *RST: 0

Example: See [Example"Trigger configuration"](#) on page 37.

Manual operation: See "[External Inhibit/Trigger Inhibit](#)" on page 27

[:SOURce<hw>]:BB:LORA:TRIGger:OBASeband:DELay <Delay>

Sets the trigger delay for triggering by the trigger signal from the other path.

Parameters:

<Delay> float
 Range: 0 to 2147483647
 Increment: 0.01
 *RST: 0
 Default unit: samples

Example: See [Example "Trigger configuration"](#) on page 37.

[[:SOURce<hw>]:BB:LORA:TRIGger:OBASeband:DElay <Delay>]

Sets the trigger delay for triggering by the trigger signal from the other path.

Parameters:

<Delay> float
 Range: 0 to 65535
 Increment: 0.01
 *RST: 0

Example: See [Example "Trigger configuration"](#) on page 37.

[[:SOURce<hw>]:BB:LORA:TRIGger:OBASeband:INHibit <Inhibit>]

For triggering via the other path, specifies the duration by which a restart is inhibited.

Parameters:

<Inhibit> integer
 Range: 0 to 67108863
 *RST: 0

Example: See [Example "Trigger configuration"](#) on page 37.

[[:SOURce<hw>]:BB:LORA:TRIGger:ARM:EXECute]

Stops signal generation; a subsequent trigger event restarts signal generation.

Example: See [Example "Trigger configuration"](#) on page 37.

Usage: Event

Manual operation: See "[Arm](#)" on page 26

[[:SOURce<hw>]:BB:LORA:TRIGger:EXECute]

Executes a trigger.

Example: See [Example "Trigger configuration"](#) on page 37.

Usage: Event

Manual operation: See "[Execute Trigger](#)" on page 26

[:SOURce<hw>]:BB:LORA:TRIGger:RMODE <RMode>

Queries the signal generation status.

Parameters:

<RMode>	STOP RUN
*RST:	STOP

Example: See [Example "Trigger configuration"](#) on page 37.

Manual operation: See ["Running/Stopped"](#) on page 25

[:SOURce<hw>]:BB:LORA:TRIGger:SLength <SLength>

Defines the length of the signal sequence that is output in the SINGLE trigger mode

Parameters:

<SLength>	integer
Range:	1 to dynamic
*RST:	1

Example: See [Example "Trigger configuration"](#) on page 37.

Manual operation: See ["Signal Duration"](#) on page 25

[:SOURce<hw>]:BB:LORA:TRIGger:SLUnit <SIUnit>

Defines the unit for the entry of the signal sequence length.

Parameters:

<SIUnit>	SEQUence SAMPLE
*RST:	SEQUence

Example:

```
SOURcel:BB:LORA:TRIGger:SEQUence SING
SOURcel:BB:LORA:TRIGger:SLUnit SEQ
SOURcel:BB:LORA:TRIGger:SLength 2
```

Manual operation: See ["Signal Duration Unit"](#) on page 24

[:SOURce<hw>]:BB:LORA:TRIGger:TIME:DATE <Year>, <Month>, <Day>

Sets the date for a time-based trigger signal. For trigger modes single or armed auto, you can activate triggering at this date via the following command:

SOURce<hw>:BB:<DigStd>:TRIGger:TIME:STATE

<DigStd> is the mnemonic for the digital standard, for example, ARB. Time-based triggering behaves analogously for all digital standards that support this feature.

Parameters:

<Year>	integer
Range:	1980 to 9999

<Month>	integer
	Range: 1 to 12
<Day>	integer
	Range: 1 to 31
Example:	See example "Configure a time-based trigger signal" in the sub-chapter "Trigger Commands" of the chapter "SOURce:BB:ARB subsystem" in the R&S SMM100A user manual.
Manual operation:	See " Trigger Time " on page 25

[[:SOURce<hw>](#)]:BB:LORA:TRIGger:TIME:TIME <Hour>, <Minute>, <Second>

Sets the time for a time-based trigger signal. For trigger modes single or armed auto, you can activate triggering at this time via the following command:

[SOURce<hw>:BB:<DigStd>:TRIGger:TIME:STATe](#)

<DigStd> is the mnemonic for the digital standard, for example, ARB. Time-based triggering behaves analogously for all digital standards that support this feature.

Parameters:

<Hour>	integer
	Range: 0 to 23
<Minute>	integer
	Range: 0 to 59
<Second>	integer
	Range: 0 to 59

Example:	See example "Configure a time-based trigger signal" in the sub-chapter "Trigger Commands" of the chapter "SOURce:BB:ARB subsystem" in the R&S SMM100A user manual.
Manual operation:	See " Trigger Time " on page 25

[[:SOURce<hw>](#)]:BB:LORA:TRIGger:TIME[:STATe] <State>

Activates time-based triggering with a fixed time reference. If activated, the R&S SMM100A triggers signal generation when its operating system time matches a specified time.

Specify the trigger date and trigger time with the following commands:

[SOURce<hw>:BB:<DigStd>:TRIGger:TIME:DATE](#)

[SOURce<hw>:BB:<DigStd>:TRIGger:TIME:TIME](#)

<DigStd> is the mnemonic for the digital standard, for example, ARB. Time-based triggering behaves analogously for all digital standards that support this feature.

Parameters:

<State> 1 | ON | 0 | OFF
 *RST: 0

Example: See example "Configure a time-based trigger signal" in the sub-chapter "Trigger Commands" of the chapter "SOURce:BB:ARB subsystem" in the R&S SMM100A user manual.

Manual operation: See "[Time Based Trigger](#)" on page 25

6.6 Marker commands

[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:MODE.....	53
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:DElay.....	53
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:ONTime.....	54
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:OFFTime.....	54
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:PATTern.....	54
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:PULSe:DIVider.....	54
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:PULSe:FREQuency?.....	54

[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:MODE <Mode>

Defines the signal for the selected marker output.

Parameters:

<Mode> RESTart | PULSe | PATTern | RATio | TRIGger | FRAMe
 *RST: RESTart

Example: See [Example"Configure and enable standard marker signals"](#) on page 37.

Manual operation: See "[Mode](#)" on page 28

[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:DElay <Delay>

Defines the delay between the signal on the marker outputs and the start of the signals.

Parameters:

<Delay> float
 Range: 0 to 16777215
 Increment: 1
 *RST: 0
 Default unit: Samples

Example: See [Example"Configure and enable standard marker signals"](#) on page 37.

Manual operation: See "[Delay](#)" on page 29

[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:ONTIme <OnTime>
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:OFFTime <OffTime>

Sets the number of samples during which the marker output is on or off.

Parameters:

<OffTime>	integer
	Range: 1 to 16777215
	*RST: 1

Example: See [Example "Configure and enable standard marker signals"](#) on page 37.

Manual operation: See "[Mode](#)" on page 28

[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:PATTern <Pattern>, <BitCount>

Sets the bit pattern defining the marker signal.

Parameters:

<Pattern>	numeric
	*RST: #H2
<BitCount>	integer
	Range: 1 to 64
	*RST: 2

Example: See [Example "Configure and enable standard marker signals"](#) on page 37.

Manual operation: See "[Mode](#)" on page 28

[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:PULSe:DIVider <Divider>

Set the pulse divider for marker mode `PULSE`.

Parameters:

<Divider>	integer
	Range: 2 to 1024
	*RST: 2

Example: See [\[:SOURce<hw>\]:BB:LORA:TRIGger:OUTPut<ch>:PULSe:FREQuency?](#) on page 54.

Manual operation: See "[Mode](#)" on page 28

[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:PULSe:FREQuency?

Queries the pulse frequency.

The return value depends on the `pulse divider`, which is the `sample rate` divided by the pulse divider.

Return values:

<Frequency> float
 Range: 2 to 1024
 Increment: 1E-3
 *RST: 2

Example:

```
SOURcel:BB:LORA:TRIGger:OUTPut1:MODE PULSe
SOURcel:BB:LORA:TRIGger:OUTPut1:PULSe:DIVider 4
SOURcel:BB:LORA:SRATE:VARiation?
// 500000
SOURcel:BB:LORA:TRIGger:OUTPut1:PULSe:FREQuency?
// 125000
```

Usage: Query only

Manual operation: See "[Mode](#)" on page 28

6.7 Clock commands

[:SOURce<hw>]:BB:LORA:CLOCK:SOURce	55
[:SOURce<hw>]:BB:LORA:CLOCK:MODE	55

[:SOURce<hw>]:BB:LORA:CLOCK:SOURce <Source>

Selects the clock source:

- INTernal: Internal clock reference
- ELClock: External local clock
- EXTernal = ELClock: Setting only

Provided for backward compatibility with other Rohde & Schwarz signal generators

Parameters:

<Source> INTernal
 *RST: INTernal

Example: See [Example "Clock settings"](#) on page 38.

Manual operation: See "[Clock Source](#)" on page 30

[:SOURce<hw>]:BB:LORA:CLOCK:MODE <Mode>

Sets the type of externally supplied clock.

Parameters:

<Mode> SAMPlE | MSAMPlE
 *RST: SAMPlE

Example: See [Example "Clock settings"](#) on page 38.

Manual operation: See "[Clock Mode](#)" on page 31

Glossary: Abbreviations and definitions

C

CEPT: European Conference of Postal and Telecommunications Administrations

CSS: Chirp Spread Spectrum

CW: Continous Wave

E

ECO: European Communications Office

ERC: European Radiocommunications Committee

ETSI: European Telecommunications Standards Institute

F

FCC: Federal Communications Commission

I

IoT: Internet of Things

L

LoRa: Longe Range

LoRaWAN: Longe Range Wide Area Network

P

PER: Packet Error Rate

PHY: Physical layer

R

REC: Physical layer

W

WAN: [LoRaWAN](#)

Glossary: Specifications and references

Symbols

1MA295: Rohde&Schwarz

Application Note 1MA295: Characterization of LoRa Devices

<https://www.rohde-schwarz.com/appnote/1MA295>

C

CEPT/ERC Recommendation 70-03: "Relating to the use of Short Range Devices (SRD)"

Available at [ECO](#) data base

<https://www.ecodocdb.dk/download/25c41779-cd6e/Rec7003e.pdf>

E

ETSI Specification EN 300 220-1: "Short Range Devices (SRD) operating in the frequency range 25 MHz to 1000 MHz"

https://www.etsi.org/deliver/etsi_en/300200_300299/30022001/

F

FCC Specification 47 Part 15.247: "47 CFR 15.247 - Operation within the bands 902-928 MHz, 2400-2483.5 MHz, and 5725-5850 MHz."

<https://www.govinfo.gov/content/pkg/CFR-2010-title47-vol1/pdf/CFR-2010-title47-vol1-sec15-247.pdf>

L

LoRaWAN Regional Parameters Specification: "LoRaWAN™ 1.1 Regional Parameters"

<https://lora-alliance.org/resource-hub/lorawantm-regional-parameters-v11ra>

LoRaWAN Specification: "LoRaWAN™ 1.1 Specification"

<https://lora-alliance.org/resource-hub/lorawantm-specification-v11>

R

R&S SGMA-GUI Software: Software for control of R&S SGS

<https://www.rohde-schwarz.com/software/sgs100a/>

R&S SGS100A user manual: <https://www.rohde-schwarz.com/manual/sgs100a/>

List of commands

[:SOURce<hw>]:BB:LORA:BWIDth.....	38
[:SOURce<hw>]:BB:LORA:CLOCK:MODE.....	55
[:SOURce<hw>]:BB:LORA:CLOCK:SOURce.....	55
[:SOURce<hw>]:BB:LORA:FCONfiguration:BMODe:STATe.....	41
[:SOURce<hw>]:BB:LORA:FCONfiguration:CMODe:STATe.....	42
[:SOURce<hw>]:BB:LORA:FCONfiguration:CRATe.....	42
[:SOURce<hw>]:BB:LORA:FCONfiguration:DATA.....	42
[:SOURce<hw>]:BB:LORA:FCONfiguration:DATA:DPATtern.....	43
[:SOURce<hw>]:BB:LORA:FCONfiguration:DATA:DSELection.....	43
[:SOURce<hw>]:BB:LORA:FCONfiguration:DLENgth.....	43
[:SOURce<hw>]:BB:LORA:FCONfiguration:EACTive:STATe.....	43
[:SOURce<hw>]:BB:LORA:FCONfiguration:HACTive:STATe.....	44
[:SOURce<hw>]:BB:LORA:FCONfiguration:IACTive:STATe.....	44
[:SOURce<hw>]:BB:LORA:FCONfiguration:PCRC:STATe.....	44
[:SOURce<hw>]:BB:LORA:FCONfiguration:PRCMode:STATe.....	44
[:SOURce<hw>]:BB:LORA:FCONfiguration:RBIT:STATe.....	45
[:SOURce<hw>]:BB:LORA:FCONfiguration:SFACtor.....	45
[:SOURce<hw>]:BB:LORA:FCONfiguration:SMODE.....	45
[:SOURce<hw>]:BB:LORA:FCONfiguration:UPLength.....	45
[:SOURce<hw>]:BB:LORA:IINTerval.....	38
[:SOURce<hw>]:BB:LORA:IMPalements:FDDeviation.....	46
[:SOURce<hw>]:BB:LORA:IMPalements:FDRate.....	46
[:SOURce<hw>]:BB:LORA:IMPalements:FDRift:STATe.....	46
[:SOURce<hw>]:BB:LORA:IMPalements:FDTYpe.....	46
[:SOURce<hw>]:BB:LORA:IMPalements:FOFFset.....	47
[:SOURce<hw>]:BB:LORA:IMPalements:STATe.....	47
[:SOURce<hw>]:BB:LORA:IMPalements:STERror.....	47
[:SOURce<hw>]:BB:LORA:OSAMpling.....	40
[:SOURce<hw>]:BB:LORA:PRESet.....	39
[:SOURce<hw>]:BB:LORA:SETTing:CATalog.....	39
[:SOURce<hw>]:BB:LORA:SETTing:DELete.....	39
[:SOURce<hw>]:BB:LORA:SETTing:LOAD.....	39
[:SOURce<hw>]:BB:LORA:SETTing:STORe.....	40
[:SOURce<hw>]:BB:LORA:SLENgth.....	40
[:SOURce<hw>]:BB:LORA:SRATE:VARiation.....	40
[:SOURce<hw>]:BB:LORA:STATe.....	41
[:SOURce<hw>]:BB:LORA:TRIGger:ARM:EXECute.....	50
[:SOURce<hw>]:BB:LORA:TRIGger:EXECute.....	50
[:SOURce<hw>]:BB:LORA:TRIGger:OBASeband:DELay.....	49
[:SOURce<hw>]:BB:LORA:TRIGger:OBASeband:DELay.....	50
[:SOURce<hw>]:BB:LORA:TRIGger:OBASeband:INHibit.....	50
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:DELay.....	53
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:MODE.....	53
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:OFFTime.....	54
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:ONTime.....	54
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:PATTern.....	54
[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:PULSe:DIVider.....	54

[:SOURce<hw>]:BB:LORA:TRIGger:OUTPut<ch>:PULSe:FREQuency?	54
[:SOURce<hw>]:BB:LORA:TRIGger:RMODE	51
[:SOURce<hw>]:BB:LORA:TRIGger:SLENgth	51
[:SOURce<hw>]:BB:LORA:TRIGger:SLUNit	51
[:SOURce<hw>]:BB:LORA:TRIGger:SOURce	48
[:SOURce<hw>]:BB:LORA:TRIGger:TIME:DATE	51
[:SOURce<hw>]:BB:LORA:TRIGger:TIME:TIME	52
[:SOURce<hw>]:BB:LORA:TRIGger:TIME[:STATe]	52
[:SOURce<hw>]:BB:LORA:TRIGger[:EXTernal]:DELay	49
[:SOURce<hw>]:BB:LORA:TRIGger[:EXTernal]:INHibit	49
[:SOURce<hw>]:BB:LORA:TRIGger[:EXTernal]:SYNChronize:OUTPut	49
[:SOURce<hw>]:BB:WAVEform:CREate	41
[:SOURce<hw>]:BB:LORA[:TRIGger]:SEQuence	48

Index

A

Application cards	7
Application notes	7
Arm	
Trigger	26
Armed	
Auto, trigger mode	24
Retrigger, trigger mode	24
Auto	
Trigger mode	24

B

Bandwidth	16
Blocking test	
Generate adjacent channel blocking signal	33
Test setup	33
Brochures	7

C

Clock	
Mode	31
Source	30
Coding	
Activate	18
Interleaver active	18
Rate	18
Conventions	
SCPI commands	35

D

Data	
Length	19
Pattern	19
Payload CRC	20
Source	19
Data sheets	7
Default settings	
LoRa	15
Delay	
Marker	29
Trigger	27
Documentation overview	6

F

Frame configuration	17
Burst mode	20
Coding rate	18
Compressed mode	20
Data	18
Data length	19
Data source	19
Encoder active	18
Header	20
Header active	20
Interleaver active	18
Modulation and coding	18
Payload CRC	20
Payload reduced coding mode	17
Reserved bit	20

Spreading factor	18
Sync mode	18
Unmodulated preamble length	17
Frequency drift	
Configuration	21
Deviation	22
Rate	22
State	21
Type	22
Frequency offset	21

G

Generate	
Waveform file	15
Getting started	6

H

Header	
Activate	20
Burst mode	20
Compressed mode	20
Reserved bit	20
Help	6
How to	
Generate adjacent channel blocking test signals	33
Generate LoRa signals	32
Generate Rx sensitivity test signals	32

I

Idle interval	16
Impairments	20
Frequency drift configuration	21
Frequency drift deviation	22
Frequency drift rate	22
Frequency drift state	21
Frequency drift type	22
Frequency offset	21
State	21
Symbol timing error	21
Installation	5
Instrument help	6
Instrument security procedures	7

L

LoRa	
About	9
Applications	13
Blocking test	33
Channel allocation	10
Frame structure	11
Generating test signals	32
Key features	13
Message structure	11
Modulation	9
LoRa testing	
Rx sensitivity test	32
LoRaWAN	11
Device communication classes	12
Network architecture	12

M

Marker delay	29
Marker mode	28
Modulation	
Spreading factor	18

O

Open source acknowledgment (OSA)	7
Oversampling	16

P

Payload reduced coding mode	17
-----------------------------------	----

R

Release notes	7
Remote control	
Programming examples	35
Retrigger	
Trigger mode	24
Rx sensitivity test	
Generate signal	32
Test setup	32

S

Safety instructions	7
Sample rate variation	16
Save/Recall	
LoRa	15
Security procedures	7
Sequence length	16
Service manual	6
Set to default	
LoRa	15
Signal duration unit	24
Signal generation status	25
Single	
Trigger	24
Standard settings	
LoRa	15
State	
LoRa	15
Symbol timing error	21
Synchronization mode	18

T

Time-based trigger	
Date	25
State	25
Time	25
Trigger	
Delay	27
External	27
External, inhibit	27
Mode	24
Signal duration	25
Source	26
Synchronize output	26

U

Unmodulated preamble length	17
User manual	6

V

Videos	7
--------------	---

W

Waveform file	
Create	15
White papers	7